

LB95546


245.1

INSTRUMENT 1) 1 CY1

Linear

μ Abs.:

34366



A= 0.0000e+000

B= 2.8833e-004

C= -2.5186e-002

Rho= 0.9998355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	-41	0.000	-41				
0.05	0.050	0.051	0.001	264	0.0 %	264				
0.2	0.200	0.185	-0.015	729	0.0 %	729				
2.5	2.500	2.498	-0.002	8752	0.0 %	8752				
5.0	5.000	5.052	0.052	17608	0.0 %	17608				
7.5	7.500	7.618	0.118	26508	0.0 %	26508				
10.0	10.000	9.883	-0.117	34366	0.0 %	34366				

LB95546 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 May 2018 18:25:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-41	-	0.0000	245.1	PPB	22 May 2018 18:26:21	S	Std
0.05 - 1	50.05	264	-	0.0500	245.1	PPB	22 May 2018 18:28:30	S	Std
0.2 - 1	50.2	729	-	0.2000	245.1	PPB	22 May 2018 18:30:38	S	Std
2.5 - 1	52.5	8752	-	2.5000	245.1	PPB	22 May 2018 18:32:46	S	Std
5.0 - 1	55	17608	-	5.0000	245.1	PPB	22 May 2018 18:34:54	S	Std
7.5 - 1	57.5	26508	-	7.5000	245.1	PPB	22 May 2018 18:37:01	S	Std
10.0 - 1	510	34366	-	10.0000	245.1	PPB	22 May 2018 18:39:10	S	Std
ICV61 - 1	ICV61	14337	4.1085	-245.1		PPB	22 May 2018 18:44:34	U	SMPL
ICB61 - 1	ICB61	-154	-0.0696	-245.1		PPB	22 May 2018 18:46:40	U	SMPL
CCV28 - 1	CCV28	17384	4.9871	-245.1		PPB	22 May 2018 18:48:47	U	SMPL
CCB28 - 1	CCB28	-180	-0.0771	-245.1		PPB	22 May 2018 18:50:53	U	SMPL
CRA - 1	CRA	250	0.0469	-245.1		PPB	22 May 2018 18:53:01	U	SMPL
HighStd - 1	HighStd	34585	9.9466	-245.1		PPB	22 May 2018 18:55:08	U	SMPL
ChkStd - 1	ChkStd	24661	7.0852	-245.1		PPB	22 May 2018 18:57:14	U	SMPL
PB109563BL - 1	PBW	-77	-0.0474	-245.1		PPB	22 May 2018 19:01:50	U	SMPL
PB109563BS - 1	LCSW	13188	3.7773	-245.1		PPB	22 May 2018 19:03:56	U	SMPL
J2908-02 - 1	EFFLUENT-METALS	121	0.0097	-245.1		PPB	22 May 2018 19:06:04	U	SMPL
J3047-01 - 1	051818-SIDI-E-D-B	-222	-0.0892	-245.1		PPB	22 May 2018 19:08:12	U	SMPL
J3047-02 - 1	051818-SIDI-E-D-M	-90	-0.0511	-245.1		PPB	22 May 2018 19:10:20	U	SMPL
J3047-03 - 1	051818-SIDI-E-D-S	-109	-0.0566	-245.1		PPB	22 May 2018 19:12:29	U	SMPL
J3047-04 - 1	051818-SIDI-E-U-B	-89	-0.0508	-245.1		PPB	22 May 2018 19:14:35	U	SMPL
CCV29 - 1	CCV29	17800	5.1070	-245.1		PPB	22 May 2018 19:16:42	U	SMPL
CCB29 - 1	CCB29	-190	-0.0800	-245.1		PPB	22 May 2018 19:18:49	U	SMPL
J3047-05 - 1	051818-SIDI-E-U-M	-88	-0.0506	-245.1		PPB	22 May 2018 19:20:56	U	SMPL
J3047-06 - 1	051818-SIDI-E-U-S	-109	-0.0566	-245.1		PPB	22 May 2018 19:23:03	U	SMPL
J3047-07 - 1	051818-JETW-E-D-B	-77	-0.0474	-245.1		PPB	22 May 2018 19:25:09	U	SMPL
J3047-08 - 1	051818-JETW-E-D-M	-91	-0.0514	-245.1		PPB	22 May 2018 19:27:16	U	SMPL
J3047-09 - 1	051818-JETW-E-D-S	-100	-0.0540	-245.1		PPB	22 May 2018 19:29:24	U	SMPL
J3047-10 - 1	051818-JETW-E-U-B	-112	-0.0575	-245.1		PPB	22 May 2018 19:31:31	U	SMPL
J3047-11 - 1	051818-JETW-E-U-M	-123	-0.0606	-245.1		PPB	22 May 2018 19:33:40	U	SMPL
J3047-12 - 1	051818-JETW-E-U-S	-113	-0.0578	-245.1		PPB	22 May 2018 19:35:48	U	SMPL
J3047-12DUP - 1	051818-JETW-E-U-SDUP	-112	-0.0575	-245.1		PPB	22 May 2018 19:37:56	U	SMPL
J3047-12MS - 1	051818-JETW-E-U-SMS	15886	4.5552	-245.1		PPB	22 May 2018 19:40:03	U	SMPL
CCV30 - 1	CCV30	17084	4.9006	-245.1		PPB	22 May 2018 19:42:10	U	SMPL
CCB30 - 1	CCB30	-208	-0.0852	-245.1		PPB	22 May 2018 19:44:17	U	SMPL
J3047-12MSD - 1	051818-JETW-E-U-SMSD	15870	4.5505	-245.1		PPB	22 May 2018 19:46:25	U	SMPL
J3047-12LX5 - 1		-184	-0.0782	-245.1		PPB	22 May 2018 19:48:32	U	SMPL
J3047-12A - 1		15806	4.5321	-245.1		PPB	22 May 2018 19:50:39	U	SMPL
CCV31 - 1	CCV31	17319	4.9683	-245.1		PPB	22 May 2018 19:52:46	U	SMPL
CCB31 - 1	CCB31	-216	-0.0875	-245.1		PPB	22 May 2018 19:54:53	U	SMPL