



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 05/23/2018

TEMP1 IN: 103 05/23/2018 07:10 TEMP OUT: 104 05/23/2018 08:11

TEMP2 IN: 104 05/23/2018 09:00 TEMP OUT: 104 05/23/2018 10:05

TEMP3 IN: 105 05/23/2018 11:30 TEMP OUT: 105 05/23/2018 13:05

TEMP4 IN: 105 05/23/2018 13:50 TEMP OUT: 105 05/23/2018 15:25

Run Number: LB95552

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95552BL	LB95552BL	1.3937	1.3937	100	1.3937	1.3937	1.3937	0.0000	0
2	LB95552BS	LB95552BS	1.3906	1.3906	100	1.4405	1.4405	1.4405	0.0499	499
3	J3066-01	052118-DRBI-E-D-B	1.3961	1.3961	200	1.4127	1.4127	1.4127	0.0166	83
4	J3066-01DUP	052118-DRBI-E-D-BDUP	1.3972	1.3972	200	1.4141	1.4141	1.4141	0.0169	84.5
5	J3066-02	052118-DRBI-E-D-M	1.3960	1.3960	200	1.4081	1.4081	1.4081	0.0121	60.5
6	J3066-03	052118-DRBI-E-D-S	1.3876	1.3876	300	1.4047	1.4047	1.4047	0.0171	57
7	J3066-04	052118-DBRI-E-U-B	1.3858	1.3858	200	1.3986	1.3986	1.3986	0.0128	64
8	J3066-05	052118-DBRI-E-U-M	1.3922	1.3922	300	1.4089	1.4089	1.4089	0.0167	55.7
9	J3066-06	052118-DBRI-E-U-S	1.3890	1.3890	300	1.4020	1.4020	1.4020	0.0130	43.3
10	J3070-01	14B-1	1.4085	1.4085	20	1.4512	1.4512	1.4512	0.0427	2135
11	J3070-02	14B-2	1.3878	1.3878	20	1.4392	1.4392	1.4392	0.0514	2570
12	J3070-03	14B-3	1.3918	1.3918	10	1.4128	1.4128	1.4128	0.0210	2100
13	J3070-04	14B-4	1.3882	1.3882	10	1.4125	1.4125	1.4125	0.0243	2430
14	J3072-01	DRAIN-WATER-TANK-1	1.4005	1.4005	50	1.4163	1.4163	1.4163	0.0158	316



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 05/23/2018TEMP1 IN: 103 05/23/2018 07:10 TEMP OUT: 104 05/23/2018 08:11TEMP2 IN: 104 05/23/2018 09:00 TEMP OUT: 104 05/23/2018 10:05TEMP3 IN: 105 05/23/2018 11:30 TEMP OUT: 105 05/23/2018 13:05TEMP4 IN: 105 05/23/2018 13:50 TEMP OUT: 105 05/23/2018 15:25Run Number: LB95552BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
A = Sample Volume (ml)						Weight (g) = C - B				
B = Empty Dish Weight (g)						Result mg/L = $\frac{D}{A} \times 1000 \times 1000$				
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C (g)										
D = Weight (g)										

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS052318

WorkList ID : 112457

Date : 5/23/2018 7:04:27 AM

1395552

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2018	Water	J3066-01 C	TSS	Cool 4 deg C	HDR103	H32	052118-DRBI-E-D-B	05/21/2018	SM2540 D
05/28/2018	Water	J3066-02 C	TSS	Cool 4 deg C	HDR103	H32	052118-DRBI-E-D-M	05/21/2018	SM2540 D
05/28/2018	Water	J3066-03 C	TSS	Cool 4 deg C	HDR103	H32	052118-DRBI-E-D-S	05/21/2018	SM2540 D
05/28/2018	Water	J3066-04 C	TSS	Cool 4 deg C	HDR103	H32	052118-DBRI-E-U-B	05/21/2018	SM2540 D
05/28/2018	Water	J3066-05 C	TSS	Cool 4 deg C	HDR103	H32	052118-DBRI-E-U-M	05/21/2018	SM2540 D
05/28/2018	Water	J3066-06 C	TSS	Cool 4 deg C	HDR103	H32	052118-DBRI-E-U-S	05/21/2018	SM2540 D
05/29/2018	Water	J3070-01 P	TSS	Cool 4 deg C	NEWY17	L31	14B-1	05/22/2018	SM2540 D
05/29/2018	Water	J3070-02 P	TSS	Cool 4 deg C	NEWY17	L31	14B-2	05/22/2018	SM2540 D
05/29/2018	Water	J3070-03 P	TSS	Cool 4 deg C	NEWY17	L31	14B-3	05/22/2018	SM2540 D
05/29/2018	Water	J3070-04 P	TSS	Cool 4 deg C	NEWY17	L31	14B-4	05/22/2018	SM2540 D
05/29/2018	Water	J3072-01 P	TSS	Cool 4 deg C	MAJO01	L52	DRAIN-WATER-TANK-1	05/22/2018	SM2540 D

Date/Time 05/23/18, 07:40
 Received by: AS
 Relinquished by: CP

Date/Time 05/23/18, 12:00
 Received by: AS
 Relinquished by: AS