



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95553

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/23/2018	08:45
2	CAL2	1	Water			21.3	7.00	05/23/2018	08:47
3	CAL3	1	Water			21.5	10.03	05/23/2018	08:49
4	ICV	1	Water			21.5	7.00	05/23/2018	08:51
5	CCV1	1	Water			21.4	2.00	05/23/2018	08:53
6	J3078-03	1	Solid	20.02	20	21.9	7.02	05/23/2018	08:55
7	J3078-03DUP	1	Solid	20.03	20	21.8	7.03	05/23/2018	08:57
8	J3078-06	1	Solid	20.04	20	21.4	7.20	05/23/2018	08:59
9	J3078-09	1	Solid	20.02	20	21.3	6.92	05/23/2018	09:01
10	J3078-12	1	Solid	20.03	20	21.1	7.03	05/23/2018	09:03
11	J3078-15	1	Solid	20.05	20	21.1	6.70	05/23/2018	09:05
12	CCV2	1	Water			21.5	12.00	05/23/2018	09:07

WORKLIST(Hardcopy Internal Chain)

LB95553

WorkList Name : corrosivity052318

WorkList ID : 112474

Date : 5/23/2018 8:22:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2018	Solid	J3078-03	Corrosivity	Cool 4 deg C	ENVI27	H31	PILE-1-1	05/22/2018	9045C
05/22/2018	Solid	J3078-06	Corrosivity	Cool 4 deg C	ENVI27	H31	PILE-1-2	05/22/2018	9045C
05/22/2018	Solid	J3078-09	Corrosivity	Cool 4 deg C	ENVI27	H31	PILE-1-3	05/22/2018	9045C
05/22/2018	Solid	J3078-12	Corrosivity	Cool 4 deg C	ENVI27	H31	PILE-1-4	05/22/2018	9045C
05/22/2018	Solid	J3078-15	Corrosivity	Cool 4 deg C	ENVI27	H31	PILE-2	05/22/2018	9045C

Date/Time 05/23/18 8:30 AM
 Received by: SP
 Relinquished by: CP

Date/Time 05/23/18 9:15 AM
 Received by: CP
 Relinquished by: SP