

LB 95555


7470A

INSTRUMENT ID: C41

Linear

μ Abs.:

37731



A= 0.0000e+000

B= 2.6214e-004

C= -2.3234e-002

Rho= 0.9997175

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	8	0.000	8				
0.2	0.200	0.183	-0.017	785	0.0 %	785				
2.5	2.500	2.496	-0.004	9610	0.0 %	9610				
5.0	5.000	5.009	0.009	19195	0.0 %	19195				
7.5	7.500	7.667	0.167	29335	0.0 %	29335				
10.0	10.000	9.868	-0.132	37731	0.0 %	37731				

LB95555 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 May 2018 10:23:14

Sample ID	Extended ID	μ Abs.	Conc.	Stdnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	23 May 2018 10:46:41	S	Std
0.2 - 1	50.2	785	-	0.2000	7470A	PPB	23 May 2018 10:48:49	S	Std
2.5 - 1	52.5	9610	-	2.5000	7470A	PPB	23 May 2018 10:50:58	S	Std
5.0 - 1	53	19195	-	5.0000	7470A	PPB	23 May 2018 10:53:06	S	Std
7.5 - 1	53.5	29335	-	7.5000	7470A	PPB	23 May 2018 10:55:14	S	Std
10.0 - 1	50	37731	-	10.0000	7470A	PPB	23 May 2018 10:57:22	S	Std
ICV62 - 1	ICV62	14985	3.9049	-7470A	PPB	23 May 2018 11:00:05	U	SMPL	
ICB62 - 1	ICB62	-163	-0.0660	-7470A	PPB	23 May 2018 11:02:11	U	SMPL	
CCV32 - 1	CCV32	19550	5.1016	-7470A	PPB	23 May 2018 11:04:17	U	SMPL	
CCB32 - 1	CCB32	-152	-0.0631	-7470A	PPB	23 May 2018 11:06:24	U	SMPL	
CRA - 1	CRA	777	0.1804	-7470A	PPB	23 May 2018 11:08:31	U	SMPL	
HighStd - 1	HighStd	38675	10.1150	-7470A	PPB	23 May 2018 11:10:38	U	SMPL	
ChkStd - 1	ChkStd	26864	7.0189	-7470A	PPB	23 May 2018 11:12:45	U	SMPL	
PB109562BL - 1	PBW	-210	-0.0783	-7470A	PPB	23 May 2018 11:14:53	U	SMPL	
PB109562BS - 1	LCSW	13614	3.5455	-7470A	PPB	23 May 2018 11:17:02	U	SMPL	
PB109562TB - 1	PB109524TB	-135	-0.0586	-7470A	PPB	23 May 2018 11:19:10	U	SMPL	
J2675-10 - 1	HR-05-052118-A	-68	-0.0411	-7470A	PPB	23 May 2018 11:21:18	U	SMPL	
J2675-10DUP - 1	HR-05-052118-ADUP	-15	-0.0272	-7470A	PPB	23 May 2018 11:23:26	U	SMPL	
J2675-10MS - 1	HR-05-052118-AMS	13374	3.4826	-7470A	PPB	23 May 2018 11:25:35	U	SMPL	
J2675-10MSD - 1	HR-05-052118-AMSD	13612	3.5450	-7470A	PPB	23 May 2018 11:27:41	U	SMPL	
CCV33 - 1	CCV33	19141	4.9944	-7470A	PPB	23 May 2018 11:29:48	U	SMPL	
CCB33 - 1	CCB33	-197	-0.0749	-7470A	PPB	23 May 2018 11:31:56	U	SMPL	
J2675-12 - 1	HR-05-052118-B	97	0.0022	-7470A	PPB	23 May 2018 11:34:03	U	SMPL	
J2679-20 - 1	EO-01-052118-A	8	-0.0211	-7470A	PPB	23 May 2018 11:36:09	U	SMPL	
J2679-22 - 1	EO-01-052118-B	24	-0.0169	-7470A	PPB	23 May 2018 11:38:16	U	SMPL	
J2679-24 - 1	EO-01-052118-C	6	-0.0217	-7470A	PPB	23 May 2018 11:40:23	U	SMPL	
J3055-04 - 1	PHASE-1	-12	-0.0264	-7470A	PPB	23 May 2018 11:42:30	U	SMPL	
J3055-08 - 1	PHASE-2	1	-0.0230	-7470A	PPB	23 May 2018 11:44:37	U	SMPL	
J3055-12 - 1	PHASE-3	7	-0.0214	-7470A	PPB	23 May 2018 11:46:45	U	SMPL	
J3056-04 - 1	TP-1	34	-0.0143	-7470A	PPB	23 May 2018 11:48:53	U	SMPL	
J3056-08 - 1	TP-2	-3	-0.0240	-7470A	PPB	23 May 2018 11:51:02	U	SMPL	
J3058-03 - 1	VNJ-229	93	0.0011	-7470A	PPB	23 May 2018 11:53:09	U	SMPL	
CCV34 - 1	CCV34	19332	5.0444	-7470A	PPB	23 May 2018 11:55:16	U	SMPL	
CCB34 - 1	CCB34	-167	-0.0670	-7470A	PPB	23 May 2018 11:57:22	U	SMPL	
J2675-10LX5 - 1		-57	-0.0382	-7470A	PPB	23 May 2018 11:59:30	U	SMPL	
J2675-10A - 1		15433	4.0224	-7470A	PPB	23 May 2018 12:01:37	U	SMPL	
CCV35 - 1	CCV35	19436	5.0717	-7470A	PPB	23 May 2018 12:03:44	U	SMPL	
CCB35 - 1	CCB35	-193	-0.0738	-7470A	PPB	23 May 2018 12:05:52	U	SMPL	