


7471B

LB 95556
INSTRUMENT ID: CV 2

Linear

 μ Abs.:

42941



A= 0.0000e+000

B= 2.3370e-004

C= -7.2440e-002

Rho= 0.9999462

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.064	-0.064	34	0.000	34				
0.2	0.200	0.214	0.014	1226	0.0 %	1226				
2.5	2.500	2.546	0.046	11204	0.0 %	11204				
5.0	5.000	5.027	0.027	21820	0.0 %	21820				
7.5	7.500	7.515	0.015	32465	0.0 %	32465				
10.0	10.000	9.963	-0.037	42941	0.0 %	42941				

LB95556 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 May 2018 12:31:17

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.00007471B	PPB		23 May 2018 12:33:45	S	Std
0.2 - 1	50.2	1226	-	0.20007471B	PPB		23 May 2018 12:36:02	S	Std
2.5 - 1	52.5	11204	-	2.50007471B	PPB		23 May 2018 12:38:20	S	Std
5.0 - 1	55	21820	-	5.00007471B	PPB		23 May 2018 12:40:37	S	Std
7.5 - 1	57.5	32465	-	7.50007471B	PPB		23 May 2018 12:42:55	S	Std
10.0 - 1	510	42941	-	10.00007471B	PPB		23 May 2018 12:47:47	S	Std
ICV63 - 1	ICV63	18312	4.2071	-7471B	PPB		23 May 2018 12:50:34	U	SMPL
ICB63 - 1	ICB63	4	-0.0715	-7471B	PPB		23 May 2018 12:55:34	U	SMPL
CCV36 - 1	CCV36	21873	5.0393	-7471B	PPB		23 May 2018 13:05:46	U	SMPL
CCB36 - 1	CCB36	116	-0.0453	-7471B	PPB		23 May 2018 13:08:02	U	SMPL
CRA - 1	CRA	1185	0.2045	-7471B	PPB		23 May 2018 13:10:17	U	SMPL
HighStd - 1	HighStd	42604	9.8841	-7471B	PPB		23 May 2018 13:12:33	U	SMPL
ChkStd - 1	ChkStd	31438	7.2746	-7471B	PPB		23 May 2018 13:14:52	U	SMPL
PB109575BL - 1	PBS	-98	-0.0953	-7471B	PPB		23 May 2018 13:17:10	U	SMPL
PB109575BS - 1	LCSS	16746	3.8411	-7471B	PPB		23 May 2018 13:19:27	U	SMPL
J2669-09 - 1	SU-02-052218	3476	0.7399	-7471B	PPB		23 May 2018 13:21:44	U	SMPL
J2923-01 - 1	LOQ-SCM-Precision-Bias-1	1021	0.1662	-7471B	PPB		23 May 2018 13:24:01	U	SMPL
J2923-02 - 1	LOQ-SCM-Precision-Bias-2	1140	0.1940	-7471B	PPB		23 May 2018 13:26:19	U	SMPL
J2923-03 - 1	LOQ-SCM-Precision-Bias-3	1132	0.1921	-7471B	PPB		23 May 2018 13:28:35	U	SMPL
J2923-04 - 1	LOQ-SCM-Precision-Bias-4	1143	0.1947	-7471B	PPB		23 May 2018 13:30:51	U	SMPL
CCV37 - 1	CCV37	21554	4.9647	-7471B	PPB		23 May 2018 13:33:07	U	SMPL
CCB37 - 1	CCB37	-71	-0.0890	-7471B	PPB		23 May 2018 13:35:23	U	SMPL
J3060-01 - 1	TP-3-WC-1	829	0.1213	-7471B	PPB		23 May 2018 13:37:40	U	SMPL
J3060-03 - 1	TP-3-WC-2	309	-0.0002	-7471B	PPB		23 May 2018 13:39:56	U	SMPL
J3060-05 - 1	TP-3-WC-3	161	-0.0348	-7471B	PPB		23 May 2018 13:42:12	U	SMPL
J3073-01 - 1	EPE-118-3(0-5)	43047	9.9877	-7471B	PPB		23 May 2018 13:44:29	U	SMPL
J3074-01 - 1	FK-BPM-02-005-ABCD	2138	0.4272	-7471B	PPB		23 May 2018 13:46:48	U	SMPL
J3074-01DUP - 1	FK-BPM-02-005-ABCD DUP	2348	0.4763	-7471B	PPB		23 May 2018 13:49:05	U	SMPL
J3074-01MS - 1	FK-BPM-02-005-ABCDMS	18670	4.2908	-7471B	PPB		23 May 2018 13:51:23	U	SMPL
J3074-01MSD - 1	FK-BPM-02-005-ABCDMSD	19435	4.4695	-7471B	PPB		23 May 2018 13:53:41	U	SMPL
J3075-01 - 1	RO-4-1	3938	0.8479	-7471B	PPB		23 May 2018 13:55:57	U	SMPL
J3078-02 - 1	PILE-1-1	1294	0.2300	-7471B	PPB		23 May 2018 13:58:14	U	SMPL
CCV38 - 1	CCV38	21538	4.9610	-7471B	PPB		23 May 2018 14:00:30	U	SMPL
CCB38 - 1	CCB38	-16	-0.0762	-7471B	PPB		23 May 2018 14:02:47	U	SMPL
J3078-05 - 1	PILE-1-2	1241	0.2176	-7471B	PPB		23 May 2018 14:05:04	U	SMPL
J3078-08 - 1	PILE-1-3	1105	0.1858	-7471B	PPB		23 May 2018 14:07:20	U	SMPL
J3078-11 - 1	PILE-1-4	1212	0.2108	-7471B	PPB		23 May 2018 14:09:37	U	SMPL
J3078-14 - 1	PILE-2	2986	0.6254	-7471B	PPB		23 May 2018 14:15:17	U	SMPL
J3074-01LX5 - 1		448	0.0323	-7471B	PPB		23 May 2018 14:17:33	U	SMPL
J3074-01A - 1		20551	4.7303	-7471B	PPB		23 May 2018 14:19:51	U	SMPL
CCV39 - 1	CCV39	21740	5.0082	-7471B	PPB		23 May 2018 14:24:44	U	SMPL
CCB39 - 1	CCB39	129	-0.0423	-7471B	PPB				