

LB95573


245.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

35447



A= 0.0000e+000

B= 2.8130e-004

C= 4.2599e-003

Rho= 0.9999780

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.044	-0.044	-170	0.000	-170				
0.05	0.050	0.045	-0.005	145	0.0 %	145				
0.2	0.200	0.220	0.020	767	0.0 %	767				
2.5	2.500	2.531	0.031	8981	0.0 %	8981				
5.0	5.000	5.003	0.003	17770	0.0 %	17770				
7.5	7.500	7.519	0.019	26716	0.0 %	26716				
10.0	10.000	9.975	-0.025	35447	0.0 %	35447				

LB95573 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 May 2018 16:47:49

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-170	-	0.0000	245.1	PPB	23 May 2018 17:20:49	S	Std
0.05 - 1	50.05	145	-	0.0500	245.1	PPB	23 May 2018 17:22:57	S	Std
0.2 - 1	50.2	767	-	0.2000	245.1	PPB	23 May 2018 17:25:06	S	Std
2.5 - 1	52.5	8981	-	2.5000	245.1	PPB	23 May 2018 17:27:14	S	Std
5.0 - 1	55	17770	-	5.0000	245.1	PPB	23 May 2018 17:29:22	S	Std
7.5 - 1	57.5	26716	-	7.5000	245.1	PPB	23 May 2018 17:31:29	S	Std
10.0 - 1	510	35447	-	10.0000	245.1	PPB	23 May 2018 17:38:54	S	Std
ICV64 - 1	ICV64	14874	4.1883	-245.1		PPB	23 May 2018 17:43:53	U	SMPL
ICB64 - 1	ICB64	-173	-0.0444	-245.1		PPB	23 May 2018 17:45:59	U	SMPL
CCV40 - 1	CCV40	17768	5.0024	-245.1		PPB	23 May 2018 17:48:06	U	SMPL
CCB40 - 1	CCB40	-225	-0.0590	-245.1		PPB	23 May 2018 17:50:13	U	SMPL
CRA - 1	CRA	136	0.0425	-245.1		PPB	23 May 2018 17:52:21	U	SMPL
HighStd - 1	HighStd	35657	10.0346	-245.1		PPB	23 May 2018 17:54:28	U	SMPL
ChkStd - 1	ChkStd	24407	6.8699	-245.1		PPB	23 May 2018 17:56:34	U	SMPL
PB109586BL - 1	PBW	-283	-0.0753	-245.1		PPB	23 May 2018 17:58:44	U	SMPL
PB109586BS - 1	LCSW	13587	3.8263	-245.1		PPB	23 May 2018 18:00:52	U	SMPL
J2923-09 - 1	LOQ-PW-Precision-Bias-1	651	0.1874	-245.1		PPB	23 May 2018 18:03:00	U	SMPL
J2923-10 - 1	LOQ-PW-Precision-Bias-2	780	0.2237	-245.1		PPB	23 May 2018 18:05:08	U	SMPL
J2923-11 - 1	LOQ-PW-Precision-Bias-3	795	0.2279	-245.1		PPB	23 May 2018 18:07:17	U	SMPL
J2923-12 - 1	LOQ-PW-Precision-Bias-4	800	0.2293	-245.1		PPB	23 May 2018 18:09:25	U	SMPL
J3066-01 - 1	052118-DRBI-E-D-B	-60	-0.0126	-245.1		PPB	23 May 2018 18:11:32	U	SMPL
CCV41 - 1	CCV41	18025	5.0747	-245.1		PPB	23 May 2018 18:13:39	U	SMPL
CCB41 - 1	CCB41	-201	-0.0523	-245.1		PPB	23 May 2018 18:15:46	U	SMPL
J3066-01DUP - 1	052118-DRBI-E-D-BDUP	-55	-0.0112	-245.1		PPB	23 May 2018 18:17:54	U	SMPL
J3066-01MS - 1	052118-DRBI-E-D-BMS	16256	4.5771	-245.1		PPB	23 May 2018 18:20:00	U	SMPL
J3066-01MSD - 1	052118-DRBI-E-D-BMSD	16241	4.5728	-245.1		PPB	23 May 2018 18:22:07	U	SMPL
J3066-02 - 1	052118-DRBI-E-D-M	-246	-0.0649	-245.1		PPB	23 May 2018 18:24:15	U	SMPL
J3066-03 - 1	052118-DRBI-E-D-S	-142	-0.0357	-245.1		PPB	23 May 2018 18:26:23	U	SMPL
J3066-04 - 1	052118-DBRI-E-U-B	-127	-0.0315	-245.1		PPB	23 May 2018 18:28:31	U	SMPL
J3066-05 - 1	052118-DBRI-E-U-M	-99	-0.0236	-245.1		PPB	23 May 2018 18:30:38	U	SMPL
J3066-06 - 1	052118-DBRI-E-U-S	-112	-0.0272	-245.1		PPB	23 May 2018 18:32:47	U	SMPL
J3070-12 - 1	14B(1-4)-COMP	241	0.0721	-245.1		PPB	23 May 2018 18:34:56	U	SMPL
CCV42 - 1	CCV42	17880	5.0339	-245.1		PPB	23 May 2018 18:37:02	U	SMPL
CCB42 - 1	CCB42	-191	-0.0495	-245.1		PPB	23 May 2018 18:39:09	U	SMPL
J3066-01LX5 - 1		-66	-0.0143	-245.1		PPB	23 May 2018 18:41:18	U	SMPL
J3066-01A - 1		15038	4.2344	-245.1		PPB	23 May 2018 18:43:25	U	SMPL
CCV43 - 1	CCV43	17509	4.9295	-245.1		PPB	23 May 2018 18:45:32	U	SMPL
CCB43 - 1	CCB43	-207	-0.0540	-245.1		PPB	23 May 2018 18:47:39	U	SMPL