



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95577

Slope : 98.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/23/2018	16:45
2	CAL2	1	Water			21.3	7.00	05/23/2018	16:47
3	CAL3	1	Water			21.4	10.03	05/23/2018	16:49
4	ICV	1	Water			21.7	7.00	05/23/2018	16:51
5	CCV1	1	Water			21.5	2.00	05/23/2018	16:53
6	J3094-05	1	Water			22.3	7.46	05/23/2018	17:35
7	J3094-05DUP	1	Water			22.4	7.47	05/23/2018	17:37
8	CCV2	1	Water			21.6	12.00	05/23/2018	17:39

LB95577

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity052318-2

WorkList ID : 112510

Date : 5/23/2018 4:13:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/23/2018	Water	J3094-05	Corrosivity	Cool 4 deg C	PSEG01	I32	HS-BASELINE-WATER	05/23/2018	9040C

Date/Time 05/23/18 4:30 PM

Received by: SP

Relinquished by: CP

Date/Time 05/23/18 5:45 PM

Received by: CP

Relinquished by: SP