


7470A

LB 95582
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37830



A= 0.0000e+000

B= 2.6310e-004

C= -1.2800e-002

Rho= 0.9999312

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-38	0.000	-38				
0.2	0.200	0.214	0.014	861	0.0 %	861				
2.5	2.500	2.486	-0.014	9499	0.0 %	9499				
5.0	5.000	5.000	0.000	19054	0.0 %	19054				
7.5	7.500	7.582	0.082	28868	0.0 %	28868				
10.0	10.000	9.940	-0.060	37830	0.0 %	37830				

LB95582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 May 2018 09:59:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-38	-	0.00007470A	PPB		24 May 2018 10:11:11	S	Std
0.2 - 1	50.2	861	-	0.20007470A	PPB		24 May 2018 10:13:20	S	Std
2.5 - 1	52.5	9499	-	2.50007470A	PPB		24 May 2018 10:15:28	S	Std
5.0 - 1	55	19054	-	5.00007470A	PPB		24 May 2018 10:17:36	S	Std
7.5 - 1	57.5	28868	-	7.50007470A	PPB		24 May 2018 10:19:44	S	Std
10.0 - 1	510	37830	-	10.00007470A	PPB		24 May 2018 10:27:03	S	Std
ICV65 - 1	ICV65	14914	3.9110	-7470A	PPB		24 May 2018 10:29:41	U	SMPL
ICB65 - 1	ICB65	-224	-0.0717	-7470A	PPB		24 May 2018 10:31:47	U	SMPL
CCV44 - 1	CCV44	18829	4.9411	-7470A	PPB		24 May 2018 10:33:54	U	SMPL
CCB44 - 1	CCB44	-223	-0.0715	-7470A	PPB		24 May 2018 10:36:01	U	SMPL
CRA - 1	CRA	867	0.2153	-7470A	PPB		24 May 2018 10:38:09	U	SMPL
HighStd - 1	HighStd	37808	9.9344	-7470A	PPB		24 May 2018 10:40:15	U	SMPL
ChkStd - 1	ChkStd	26805	7.0395	-7470A	PPB		24 May 2018 10:42:22	U	SMPL
PB109584BL - 1	PBW	-229	-0.0730	-7470A	PPB		24 May 2018 10:44:31	U	SMPL
PB109584BS - 1	LCSW	15025	3.9402	-7470A	PPB		24 May 2018 10:46:40	U	SMPL
PB109584TB - 1	PB109573TB	-133	-0.0478	-7470A	PPB		24 May 2018 10:48:47	U	SMPL
J2669-10 - 1	SU-02-052218	-16	-0.0170	-7470A	PPB		24 May 2018 10:50:56	U	SMPL
J3060-02 - 1	TP-3-WC-1	-36	-0.0223	-7470A	PPB		24 May 2018 10:53:04	U	SMPL
J3060-04 - 1	TP-3-WC-2	2	-0.0123	-7470A	PPB		24 May 2018 10:55:13	U	SMPL
J3060-06 - 1	TP-3-WC-3	-55	-0.0273	-7470A	PPB		24 May 2018 10:57:19	U	SMPL
CCV45 - 1	CCV45	19211	5.0416	-7470A	PPB		24 May 2018 10:59:26	U	SMPL
CCB45 - 1	CCB45	-193	-0.0636	-7470A	PPB		24 May 2018 11:01:32	U	SMPL
J3062-01 - 1	AGFA-CURIX-MR-400-CX	542	0.1298	-7470A	PPB		24 May 2018 11:03:39	U	SMPL
J3062-02 - 1	AGFA-CURIX-ULTRA-UV-L	-9	-0.0152	-7470A	PPB		24 May 2018 11:05:46	U	SMPL
J3062-03 - 1	OUPONT-CRONEX-OUANTA-II	2519	0.6499	-7470A	PPB		24 May 2018 11:07:52	U	SMPL
J3071-01 - 1	SOIL-CUTTINGS	-52	-0.0265	-7470A	PPB		24 May 2018 11:09:59	U	SMPL
J3075-04 - 1	RO-4-1	-30	-0.0207	-7470A	PPB		24 May 2018 11:12:07	U	SMPL
J3078-03 - 1	PILE-1-1	-18	-0.0175	-7470A	PPB		24 May 2018 11:14:15	U	SMPL
J3078-06 - 1	PILE-1-2	-24	-0.0191	-7470A	PPB		24 May 2018 11:16:23	U	SMPL
J3078-09 - 1	PILE-1-3	-25	-0.0194	-7470A	PPB		24 May 2018 11:18:31	U	SMPL
J3078-12 - 1	PILE-1-4	-45	-0.0246	-7470A	PPB		24 May 2018 11:20:40	U	SMPL
J3078-12DUP - 1	PILE-1-4DUP	-38	-0.0228	-7470A	PPB		24 May 2018 11:22:47	U	SMPL
CCV46 - 1	CCV46	18255	4.7900	-7470A	PPB		24 May 2018 11:24:54	U	SMPL
CCB46 - 1	CCB46	-158	-0.0544	-7470A	PPB		24 May 2018 11:27:00	U	SMPL
J3078-12MS - 1	PILE-1-4MS	13688	3.5885	-7470A	PPB		24 May 2018 11:29:08	U	SMPL
J3078-12MSD - 1	PILE-1-4MSD	13881	3.6393	-7470A	PPB		24 May 2018 11:31:14	U	SMPL
PB109585BL - 1	PBW	-136	-0.0486	-7470A	PPB		24 May 2018 11:33:22	U	SMPL
PB109585BS - 1	LCSW	14348	3.7621	-7470A	PPB		24 May 2018 11:35:29	U	SMPL
J2923-05 - 1	LOQ-NPW-Precision-Bias-1	670	0.1635	-7470A	PPB		24 May 2018 11:37:36	U	SMPL
J2923-06 - 1	LOQ-NPW-Precision-Bias-2	800	0.1977	-7470A	PPB		24 May 2018 11:39:44	U	SMPL
J2923-07 - 1	LOQ-NPW-Precision-Bias-3	791	0.1953	-7470A	PPB		24 May 2018 11:41:52	U	SMPL
J2923-08 - 1	LOQ-NPW-Precision-Bias-4	827	0.2048	-7470A	PPB		24 May 2018 11:44:00	U	SMPL
J3078-12LX5 - 1		-46	-0.0249	-7470A	PPB		24 May 2018 11:46:09	U	SMPL
J3078-12A - 1		14536	3.8116	-7470A	PPB		24 May 2018 11:48:16	U	SMPL
CCV47 - 1	CCV47	18905	4.9611	-7470A	PPB		24 May 2018 11:50:23	U	SMPL
CCB47 - 1	CCB47	-210	-0.0681	-7470A	PPB		24 May 2018 11:52:31	U	SMPL