



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95585

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	05/24/2018	09:15
2	J3094-05		1	>100.0	>212.0	>212.0	05/24/2018	09:50
3	J3094-05DUP		1	>100.0	>212.0	>212.0	05/24/2018	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

WORKLIST(Hardcopy Internal Chain)

LB95585

WorkList Name : FP-052418

WorkList ID : 112542

Date : 5/24/2018 8:18:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/27/2018	Water	J3094-05	Flash Point	Cool 4 deg C	PSEG01	I32	HS-BASELINE-WATER	05/23/2018	1010A

Date/Time 05/24/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time 05/24/18 10:25 AM

Received by: CP

Relinquished by: KS