


LB 95596
INSTRUMENT ID: LV1

7471B

Linear

μ Abs.:

47710



A= 0.0000e+000

B= 2.0952e-004

C= 2.5520e-003

Rho= 0.9999517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	27	0.000	27				
0.2	0.200	0.215	0.015	1013	0.0 %	1013				
2.5	2.500	2.438	-0.062	11623	0.0 %	11623				
5.0	5.000	5.058	0.058	24131	0.0 %	24131				
7.5	7.500	7.482	-0.018	35700	0.0 %	35700				
10.0	10.000	9.999	-0.001	47710	0.0 %	47710				

LB95596 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2018 14:32:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	27	-	0.0000	7471B	PPB	24 May 2018 14:37:40	S	Std
0.2 - 1	50.2	1013	-	0.2000	7471B	PPB	24 May 2018 14:39:58	S	Std
2.5 - 1	52.5	11623	-	2.5000	7471B	PPB	24 May 2018 14:42:15	S	Std
5.0 - 1	55	24131	-	5.0000	7471B	PPB	24 May 2018 14:44:32	S	Std
7.5 - 1	57.5	35700	-	7.5000	7471B	PPB	24 May 2018 14:46:50	S	Std
10.0 - 1	510	47710	-	10.0000	7471B	PPB	24 May 2018 14:51:47	S	Std
ICV66 - 1	ICV66	18259	3.8281	-7471B	PPB	PPB	24 May 2018 14:54:31	U	SMPL
ICB66 - 1	ICB66	6	0.0038	-7471B	PPB	PPB	24 May 2018 14:59:18	U	SMPL
CCV51 - 1	CCV51	24271	5.0877	-7471B	PPB	PPB	24 May 2018 15:01:34	U	SMPL
CCB51 - 1	CCB51	-145	-0.0278	-7471B	PPB	PPB	24 May 2018 15:03:49	U	SMPL
CRA - 1	CRA	1010	0.2142	-7471B	PPB	PPB	24 May 2018 15:06:06	U	SMPL
HighStd - 1	HighStd	47906	10.0396	-7471B	PPB	PPB	24 May 2018 15:08:22	U	SMPL
ChkStd - 1	ChkStd	30657	6.4257	-7471B	PPB	PPB	24 May 2018 15:10:38	U	SMPL
PB109618BL - 1	PBS	-125	-0.0236	-7471B	PPB	PPB	24 May 2018 15:12:56	U	SMPL
PB109618BS - 1	LCSS	16447	3.4485	-7471B	PPB	PPB	24 May 2018 15:15:15	U	SMPL
J2670-05 - 1	TR-04-052318	11457	2.4030	-7471B	PPB	PPB	24 May 2018 15:17:32	U	SMPL
J2672-07 - 1	BU-02-052318	1495	0.3158	-7471B	PPB	PPB	24 May 2018 15:19:49	U	SMPL
J2678-03 - 1	OR-03-052318	1814	0.3826	-7471B	PPB	PPB	24 May 2018 15:22:07	U	SMPL
J3084-01 - 1	RO-4-1-A	23022	4.8260	-7471B	PPB	PPB	24 May 2018 15:24:25	U	SMPL
J3084-05 - 1	RO-4-1-B	39183	8.2120	-7471B	PPB	PPB	24 May 2018 15:26:41	U	SMPL
CCV52 - 1	CCV52	23973	5.0253	-7471B	PPB	PPB	24 May 2018 15:28:58	U	SMPL
CCB52 - 1	CCB52	-153	-0.0295	-7471B	PPB	PPB	24 May 2018 15:31:16	U	SMPL
J3084-09 - 1	RO-4-1-C	35187	7.3748	-7471B	PPB	PPB	24 May 2018 15:33:32	U	SMPL
J3086-01 - 1	72-12013	13814	2.8968	-7471B	PPB	PPB	24 May 2018 15:35:48	U	SMPL
J3094-01 - 1	HS-RO-7212003	772	0.1643	-7471B	PPB	PPB	24 May 2018 15:38:06	U	SMPL
J3095-01 - 1	TP-3-WC-4	434	0.0935	-7471B	PPB	PPB	24 May 2018 15:40:22	U	SMPL
J3096-01 - 1	FORT-1-E(0-1)	4951	1.0399	-7471B	PPB	PPB	24 May 2018 15:42:39	U	SMPL
J3096-02 - 1	FORT-1-E(4-5)	519	0.1113	-7471B	PPB	PPB	24 May 2018 15:44:56	U	SMPL
J3096-03 - 1	FORT-1-E(7-8)	403	0.0870	-7471B	PPB	PPB	24 May 2018 15:47:13	U	SMPL
J3096-03DUP - 1	FORT-1-E(7-8)DUP	366	0.0792	-7471B	PPB	PPB	24 May 2018 15:49:31	U	SMPL
J3096-03MS - 1	FORT-1-E(7-8)MS	18522	3.8832	-7471B	PPB	PPB	24 May 2018 15:51:49	U	SMPL
J3096-03MSD - 1	FORT-1-E(7-8)MSD	18032	3.7805	-7471B	PPB	PPB	24 May 2018 15:54:05	U	SMPL
CCV53 - 1	CCV53	24229	5.0789	-7471B	PPB	PPB	24 May 2018 15:56:21	U	SMPL
CCB53 - 1	CCB53	0	0.0026	-7471B	PPB	PPB	24 May 2018 16:08:03	U	SMPL
PB109619BL - 1	PBS	80	0.0193	-7471B	PPB	PPB	24 May 2018 16:10:18	U	SMPL
PB109619BS - 1	LCSS	16861	3.5352	-7471B	PPB	PPB	24 May 2018 16:12:34	U	SMPL
J3098-01 - 1	EPE-F-3A(35-37)	563	0.1205	-7471B	PPB	PPB	24 May 2018 16:14:51	U	SMPL
J3098-02 - 1	EPE-F-3A(55-57)	504	0.1081	-7471B	PPB	PPB	24 May 2018 16:17:07	U	SMPL
J3098-03 - 1	EPE-F-3A(70-72)	1175	0.2487	-7471B	PPB	PPB	24 May 2018 16:19:24	U	SMPL
J3099-01 - 1	200028	8171	1.7145	-7471B	PPB	PPB	24 May 2018 16:21:41	U	SMPL
J3099-01DUP - 1	200028DUP	7293	1.5306	-7471B	PPB	PPB	24 May 2018 16:23:58	U	SMPL
J3099-01MS - 1	200028MS	24535	5.1430	-7471B	PPB	PPB	24 May 2018 16:26:16	U	SMPL
J3099-01MSD - 1	200028MSD	26304	5.5137	-7471B	PPB	PPB	24 May 2018 16:28:34	U	SMPL
J3099-04 - 1	200027	19747	4.1399	-7471B	PPB	PPB	24 May 2018 16:30:51	U	SMPL
CCV54 - 1	CCV54	24347	5.1036	-7471B	PPB	PPB	24 May 2018 16:33:08	U	SMPL
CCB54 - 1	CCB54	379	0.0820	-7471B	PPB	PPB	24 May 2018 16:35:25	U	SMPL
J3100-01 - 1	DRUM-1-A	1402	0.2963	-7471B	PPB	PPB	24 May 2018 16:37:41	U	SMPL
J3100-03 - 1	DRUM-1-B	2674	0.5628	-7471B	PPB	PPB	24 May 2018 16:39:57	U	SMPL
J3101-01 - 1	FK-PS-01-027	553	0.1184	-7471B	PPB	PPB	24 May 2018 16:42:13	U	SMPL
J3104-01 - 1	SP-A	1222	0.2586	-7471B	PPB	PPB	24 May 2018 16:44:30	U	SMPL
J3104-03 - 1	SP-B	1689	0.3564	-7471B	PPB	PPB	24 May 2018 16:46:47	U	SMPL
J3096-03LX5 - 1		73	0.0178	-7471B	PPB	PPB	24 May 2018 16:49:03	U	SMPL
J3096-03A - 1		18482	3.8748	-7471B	PPB	PPB	24 May 2018 16:51:20	U	SMPL
J3099-01LX5 - 1		1497	0.3162	-7471B	PPB	PPB	24 May 2018 16:53:38	U	SMPL
J3099-01A - 1		25825	5.4133	-7471B	PPB	PPB	24 May 2018 16:55:56	U	SMPL
CCV55 - 1	CCV55	23651	4.9578	-7471B	PPB	PPB	24 May 2018 16:58:12	U	SMPL
CCB55 - 1	CCB55	0	0.0026	-7471B	PPB	PPB	24 May 2018 17:03:06	U	SMPL