



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95598

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/24/2018	13:40
2	CAL2	1	Water			21.4	7.00	05/24/2018	13:42
3	CAL3	1	Water			21.3	10.03	05/24/2018	13:44
4	ICV	1	Water			21.5	7.00	05/24/2018	13:46
5	CCV1	1	Water			21.6	2.00	05/24/2018	13:48
6	J3116-01	1	Water			20.3	11.58	05/24/2018	14:00
7	J3116-01DUP	1	Water			20.4	11.59	05/24/2018	14:02
8	J3125-01	1	Water			21.2	7.38	05/24/2018	14:04
9	CCV2	1	Water			21.4	12.00	05/24/2018	14:06

WORKLIST(Hardcopy Internal Chain)

LB 95598

WorkList Name : pH052418

WorkList ID : 112569

Date : 5/24/2018 1:26:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2018	Water	J3116-01	pH	Cool 4 deg C	PSEG01	O12	ETS	05/24/2018	9040C
05/24/2018	Water	J3125-01	pH	Cool 4 deg C	PSEG01	O13	72-12013	05/24/2018	9040C

Date/Time 05/24/18 1:30 PM

Received by: SP

Relinquished by: CJP

Date/Time 05/24/18 2:15 PM

Received by: CJP

Relinquished by: SP