



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95607

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/24/2018	13:40
2	CAL2	1	Water			21.4	7.00	05/24/2018	13:42
3	CAL3	1	Water			21.3	10.03	05/24/2018	13:44
4	ICV	1	Water			21.5	7.00	05/24/2018	13:46
5	CCV1	1	Water			21.4	2.00	05/24/2018	15:20
6	J2679-26	1	Solid	20.02	20	22.6	7.58	05/24/2018	15:22
7	J2679-26DUP	1	Solid	20.03	20	22.7	7.59	05/24/2018	15:24
8	J2679-28	1	Solid	20.05	20	22.9	8.36	05/24/2018	15:26
9	J2679-30	1	Solid	20.02	20	23.2	8.65	05/24/2018	15:28
10	J3114-03	1	Solid	20.04	20	21.5	7.85	05/24/2018	15:30
11	J3114-07	1	Solid	20.02	20	21.0	7.87	05/24/2018	15:32
12	J3128-03	1	Solid	20.03	20	20.9	5.76	05/24/2018	15:34
13	J3128-06	1	Solid	20.04	20	20.5	9.22	05/24/2018	15:36
14	J3130-02	1	Solid	20.02	20	22.0	7.43	05/24/2018	15:38
15	J3130-04	1	Solid	20.03	20	21.4	6.06	05/24/2018	15:40
16	CCV2	1	Water			21.5	12.00	05/24/2018	15:42
17	J3130-06	1	Solid	20.02	20	21.3	6.58	05/24/2018	15:44
18	J3130-08	1	Solid	20.04	20	21.3	8.21	05/24/2018	15:46
19	CCV3	1	Water			21.6	2.00	05/24/2018	15:48

LB95607

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity052418-1

WorkList ID : 112573

Date : 5/24/2018 3:00:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/24/2018	Solid	J2679-26	Corrosivity	Cool 4 deg C	PSEG05		EO-03-052418-A	05/24/2018	9045C
05/24/2018	Solid	J2679-28	Corrosivity	Cool 4 deg C	PSEG05		EO-03-052418-B	05/24/2018	9045C
05/24/2018	Solid	J2679-30	Corrosivity	Cool 4 deg C	PSEG05		EO-03-052418-C	05/24/2018	9045C
05/24/2018	Solid	J3114-03	Corrosivity	Cool 4 deg C	PSEG01	O11	NYS-RT-3460	05/24/2018	9045C
05/24/2018	Solid	J3114-07	Corrosivity	Cool 4 deg C	PSEG01	O11	NYS-RBR-200035	05/24/2018	9045C
05/24/2018	Solid	J3128-03	Corrosivity	Cool 4 deg C	PSEG01	I13	TJ-SOIL-RO-72-11995	05/24/2018	9045C
05/24/2018	Solid	J3128-06	Corrosivity	Cool 4 deg C	PSEG01	I13	TJ-OILY-GRAVEL	05/24/2018	9045C
05/24/2018	Solid	J3130-02	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-WC-1	05/24/2018	9045C
05/24/2018	Solid	J3130-04	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-WC-2	05/24/2018	9045C
05/24/2018	Solid	J3130-06	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-WC-3	05/24/2018	9045C
05/24/2018	Solid	J3130-08	Corrosivity	Cool 4 deg C	PSEG01	O21	TP-5-WC-4	05/24/2018	9045C

Date/Time 05/24/18 3:15 PM

Received by: SP

Relinquished by: CP

Date/Time 05/24/18 4:10 PM

Received by: CP

Relinquished by: SP