

ISM02.4


LB95011

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44097



A= 0.0000e+000

B= 2.2630e-004

C= -1.8550e-002

Rho= 0.9998855

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.007	-0.007	49	0.000	49					
0.2	0.200	0.225	0.025	1076	0.0 %	1076					
2.5	2.500	2.498	-0.002	11119	0.0 %	11119					
5.0	5.000	4.922	-0.078	21833	0.0 %	21833					
7.5	7.500	7.602	0.102	33675	0.0 %	33675					
10.0	10.000	9.961	-0.039	44097	0.0 %	44097					

103
5/24/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	49	-		0 ISM02.4	PPB	5/24/2018 17:21	Std
	0.2 S0.2	1076	-		0.2 ISM02.4	PPB	5/24/2018 17:23	Std
	2.5 S2.5	11119	-		2.5 ISM02.4	PPB	5/24/2018 17:25	Std
	5 S5	21833	-		5 ISM02.4	PPB	5/24/2018 17:28	Std
	7.5 S7.5	33675	-		7.5 ISM02.4	PPB	5/24/2018 17:34	Std
	10 S10.0	44097	-		10 ISM02.4	PPB	5/24/2018 17:36	Std
ICV	ICV	16880	3.8014 -		ISM02.4	PPB	5/24/2018 17:39	SMPL
ICB	ICB	-125	-0.0468 -		ISM02.4	PPB	5/24/2018 17:41	SMPL
CCV056	CCV056	21639	4.8783 -		ISM02.4	PPB	5/24/2018 17:44	SMPL
CCB056	CCB056	-127	-0.0473 -		ISM02.4	PPB	5/24/2018 17:46	SMPL
PB109649BL	PBW003	26	-0.0127 -		ISM02.4	PPB	5/24/2018 17:48	SMPL
J3069-01	MH1C72	-18	-0.0226 -		ISM02.4	PPB	5/24/2018 17:51	SMPL
J3069-02	MH1C73	-72	-0.0348 -		ISM02.4	PPB	5/24/2018 17:53	SMPL
J3117-01	MC0AA6	-5	-0.0197 -		ISM02.4	PPB	5/24/2018 17:55	SMPL
J3117-02	MC0AA7	-41	-0.0278 -		ISM02.4	PPB	5/24/2018 17:57	SMPL
J3117-03	MC0AA8	-86	-0.038 -		ISM02.4	PPB	5/24/2018 18:00	SMPL
J3117-04	MC0AA9	5	-0.0174 -		ISM02.4	PPB	5/24/2018 18:02	SMPL
J3117-05	MC0AB0	-102	-0.0416 -		ISM02.4	PPB	5/24/2018 18:04	SMPL
J3117-06	MC0AB1	-109	-0.0432 -		ISM02.4	PPB	5/24/2018 18:07	SMPL
J3117-07	MC0AA6D	-111	-0.0437 -		ISM02.4	PPB	5/24/2018 18:09	SMPL
J3117-08	MC0AA6S	4420	0.9817 -		ISM02.4	PPB	5/24/2018 18:11	SMPL
CCV057	CCV057	21801	4.915 -		ISM02.4	PPB	5/24/2018 18:13	SMPL
CCB057	CCB057	-111	-0.0437 -		ISM02.4	PPB	5/24/2018 18:16	SMPL