


7470A

LB 95629
INSTRUMENT ID: CV1

Linear

 μ Abs.:

37604



A= 0.0000e+000

B= 2.6705e-004

C= 1.0229e-002

Rho= 0.9999431

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-6	0.000	-6				
0.2	0.200	0.223	0.023	795	0.0 %	795				
2.5	2.500	2.517	0.017	9385	0.0 %	9385				
5.0	5.000	4.931	-0.069	18428	0.0 %	18428				
7.5	7.500	7.468	-0.032	27928	0.0 %	27928				
10.0	10.000	10.052	0.052	37604	0.0 %	37604				

LB95629 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 May 2018 12:40:23

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.00007470A	PPB	25 May 2018 12:42:05	S	Std
0.2 - 1	50.2	795	-	0.20007470A	PPB	25 May 2018 12:44:14	S	Std
2.5 - 1	52.5	9385	-	2.50007470A	PPB	25 May 2018 12:46:23	S	Std
5.0 - 1	55	18428	-	5.00007470A	PPB	25 May 2018 12:48:31	S	Std
7.5 - 1	57.5	27928	-	7.50007470A	PPB	25 May 2018 12:50:39	S	Std
10.0 - 1	511	37604	-	10.00007470A	PPB	25 May 2018 12:55:16	S	Std
ICV68 - 1	ICV68	15737	4.2128	-7470A	PPB	25 May 2018 12:57:48	U	SMPL
ICB68 - 1	ICB68	-226	-0.0501	-7470A	PPB	25 May 2018 12:59:54	U	SMPL
CCV64 - 1	CCV64	18619	4.9825	-7470A	PPB	25 May 2018 13:02:02	U	SMPL
CCB64 - 1	CCB64	-223	-0.0493	-7470A	PPB	25 May 2018 13:04:08	U	SMPL
CRA - 1	CRA	773	0.2167	-7470A	PPB	25 May 2018 13:06:15	U	SMPL
HighStd - 1	HighStd	38230	10.2196	-7470A	PPB	25 May 2018 13:08:22	U	SMPL
ChkStd - 1	ChkStd	25667	6.8646	-7470A	PPB	25 May 2018 13:10:29	U	SMPL
PB109658BL - 1	PBW	-276	-0.0635	-7470A	PPB	25 May 2018 13:12:38	U	SMPL
PB109658BS - 1	LCSW	14119	3.7807	-7470A	PPB	25 May 2018 13:14:48	U	SMPL
PB109658TB - 1	PB109614TB	-173	-0.0360	-7470A	PPB	25 May 2018 13:16:55	U	SMPL
J2670-06 - 1	TR-04-052318	-64	-0.0069	-7470A	PPB	25 May 2018 13:19:03	U	SMPL
J2670-06DUP - 1	TR-04-052318DUP	-65	-0.0071	-7470A	PPB	25 May 2018 13:21:11	U	SMPL
J2670-06MS - 1	TR-04-052318MS	14512	3.8857	-7470A	PPB	25 May 2018 13:23:20	U	SMPL
J2670-06MSD - 1	TR-04-052318MSD	14309	3.8315	-7470A	PPB	25 May 2018 13:25:26	U	SMPL
CCV65 - 1	CCV65	18670	4.9961	-7470A	PPB	25 May 2018 13:27:33	U	SMPL
CCB65 - 1	CCB65	-213	-0.0467	-7470A	PPB	25 May 2018 13:29:41	U	SMPL
J2672-08 - 1	BU-02-052318	-67	-0.0077	-7470A	PPB	25 May 2018 13:31:48	U	SMPL
J2678-04 - 1	OR-03-052318	-43	-0.0013	-7470A	PPB	25 May 2018 13:33:55	U	SMPL
J3077-01 - 1	CS-RO-7211969	-15	0.0062	-7470A	PPB	25 May 2018 13:36:02	U	SMPL
J3084-04 - 1	RO-4-1-A	-16	0.0060	-7470A	PPB	25 May 2018 13:38:09	U	SMPL
J3084-08 - 1	RO-4-1-B	-38	0.0001	-7470A	PPB	25 May 2018 13:40:16	U	SMPL
J3084-12 - 1	RO-4-1-C	-23	0.0041	-7470A	PPB	25 May 2018 13:42:24	U	SMPL
J3086-02 - 1	72-12013	-10	0.0076	-7470A	PPB	25 May 2018 13:44:32	U	SMPL
J3094-03 - 1	HS-RO-7212003	-20	0.0049	-7470A	PPB	25 May 2018 13:46:40	U	SMPL
J3095-02 - 1	TP-3-WC-4	-28	0.0028	-7470A	PPB	25 May 2018 13:48:49	U	SMPL
J3099-03 - 1	200028	-40	-0.0005	-7470A	PPB	25 May 2018 13:50:56	U	SMPL
CCV66 - 1	CCV66	18587	4.9739	-7470A	PPB	25 May 2018 13:53:03	U	SMPL
CCB66 - 1	CCB66	-192	-0.0410	-7470A	PPB	25 May 2018 13:55:10	U	SMPL
J3099-06 - 1	200027	-49	-0.0029	-7470A	PPB	25 May 2018 13:57:18	U	SMPL
J3100-02 - 1	DRUM-1-A	-33	0.0014	-7470A	PPB	25 May 2018 13:59:25	U	SMPL
J3100-04 - 1	DRUM-1-B	32	0.0188	-7470A	PPB	25 May 2018 14:01:32	U	SMPL
J3104-02 - 1	SP-A	-34	0.0011	-7470A	PPB	25 May 2018 14:03:39	U	SMPL
J3104-04 - 1	SP-B	-42	-0.0010	-7470A	PPB	25 May 2018 14:05:46	U	SMPL
PB109659BL - 1	PBW	-18	0.0054	-7470A	PPB	25 May 2018 14:07:54	U	SMPL
PB109659BS - 1	LCSW	14139	3.7861	-7470A	PPB	25 May 2018 14:10:02	U	SMPL
PB109659TB - 1	PB109616TB	-135	-0.0258	-7470A	PPB	25 May 2018 14:12:10	U	SMPL
J3094-05 - 1	HS-BASELINE-WATER	-44	-0.0015	-7470A	PPB	25 May 2018 14:14:18	U	SMPL
J3094-05DUP - 1	HS-BASELINE-WATERDUP	3	0.0110	-7470A	PPB	25 May 2018 14:16:25	U	SMPL
CCV67 - 1	CCV67	18678	4.9982	-7470A	PPB	25 May 2018 14:18:33	U	SMPL
CCB67 - 1	CCB67	-198	-0.0426	-7470A	PPB	25 May 2018 14:20:40	U	SMPL
J3094-05MS - 1	HS-BASELINE-WATERMS	14192	3.8002	-7470A	PPB	25 May 2018 14:22:48	U	SMPL
J3094-05MSD - 1	HS-BASELINE-WATERMSD	14711	3.9388	-7470A	PPB	25 May 2018 14:24:55	U	SMPL
PB109673BL - 1	PBW	-168	-0.0346	-7470A	PPB	25 May 2018 14:27:02	U	SMPL
PB109673BS - 1	LCSW	15212	4.0726	-7470A	PPB	25 May 2018 14:29:09	U	SMPL
J3094-04 - 1	HS-BASELINE-WATER	-60	-0.0058	-7470A	PPB	25 May 2018 14:31:16	U	SMPL
J3094-04DUP - 1	HS-BASELINE-WATERDUP	-70	-0.0085	-7470A	PPB	25 May 2018 14:33:24	U	SMPL
J3094-04MS - 1	HS-BASELINE-WATERMS	15411	4.1258	-7470A	PPB	25 May 2018 14:35:32	U	SMPL
J3094-04MSD - 1	HS-BASELINE-WATERMSD	15394	4.1212	-7470A	PPB	25 May 2018 14:37:40	U	SMPL
J3116-01 - 1	ETS	-220	-0.0485	-7470A	PPB	25 May 2018 14:39:49	U	SMPL
J3125-01 - 1	72-12013	39169	10.4704	-7470A	PPB	25 May 2018 14:41:57	U	SMPL
CCV68 - 1	CCV68	18289	4.8943	-7470A	PPB	25 May 2018 14:44:04	U	SMPL
CCB68 - 1	CCB68	-193	-0.0413	-7470A	PPB	25 May 2018 14:46:14	U	SMPL
J3135-01 - 1	PZ-01	19093	5.1090	-7470A	PPB	25 May 2018 14:48:22	U	SMPL
J3135-02 - 1	PZ-02	23106	6.1807	-7470A	PPB	25 May 2018 14:50:29	U	SMPL
J3135-03 - 1	PZ-15	496	0.1427	-7470A	PPB	25 May 2018 14:52:38	U	SMPL
J3135-04 - 1	GMW-10	281	0.0853	-7470A	PPB	25 May 2018 14:54:47	U	SMPL
J2670-06LX5 - 1		-49	-0.0029	-7470A	PPB	25 May 2018 14:56:54	U	SMPL
J2670-06A - 1		15004	4.0171	-7470A	PPB	25 May 2018 15:01:29	U	SMPL
J3094-05LX5 - 1		-127	-0.0237	-7470A	PPB	25 May 2018 15:03:37	U	SMPL
J3094-05A - 1		15144	4.0545	-7470A	PPB	25 May 2018 15:05:45	U	SMPL
J3094-04LX5 - 1		-141	-0.0274	-7470A	PPB	25 May 2018 15:07:53	U	SMPL
J3094-04A - 1		14982	4.0112	-7470A	PPB	25 May 2018 15:10:01	U	SMPL
CCV69 - 1	CCV69	18371	4.9162	-7470A	PPB	25 May 2018 15:12:08	U	SMPL
CCB69 - 1	CCB69	-199	-0.0429	-7470A	PPB	25 May 2018 15:14:16	U	SMPL
J3125-01DLX2 - 1		18927	5.0647	-7470A	PPB	25 May 2018 15:16:24	U	SMPL
CCV70 - 1	CCV70	18219	4.8756	-7470A	PPB	25 May 2018 15:18:31	U	SMPL
CCB70 - 1	CCB70	-166	-0.0341	-7470A	PPB	25 May 2018 15:20:40	U	SMPL