



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95652

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	05/25/2018	14:00
2	CAL2	1	Water			21.5	7.00	05/25/2018	14:02
3	CAL3	1	Water			21.6	10.03	05/25/2018	14:04
4	ICV	1	Water			21.4	7.00	05/25/2018	14:06
5	CCV1	1	Water			21.5	2.00	05/25/2018	14:08
6	J3139-03	1	Water			21.6	6.76	05/25/2018	14:10
7	J3139-03DUP	1	Water			21.7	6.77	05/25/2018	14:12
8	CCV2	1	Water			21.5	12.00	05/25/2018	14:14