



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB95660

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/27/2018	08:45
2	CAL2	1	Water			21.3	7.00	05/27/2018	08:47
3	CAL3	1	Water			21.5	10.03	05/27/2018	08:49
4	ICV	1	Water			21.6	7.00	05/27/2018	08:51
5	CCV1	1	Water			21.4	2.00	05/27/2018	08:53
6	J3169-01	1	Water			21.0	6.96	05/27/2018	08:55
7	J3169-01DUP	1	Water			21.1	6.97	05/27/2018	08:57
8	J3169-04	1	Water			20.3	6.78	05/27/2018	08:59
9	J3169-07	1	Water			20.5	9.76	05/27/2018	09:01
10	J3169-10	1	Water			21.2	6.07	05/27/2018	09:03
11	J3169-14	1	Water			21.4	6.64	05/27/2018	09:05
12	CCV2	1	Water			21.6	12.00	05/27/2018	09:07

WORKLIST(Hardcopy Internal Chain)

LB95660

WorkList Name : pH052718

WorkList ID : 112655

Date : 5/27/2018 8:38:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/25/2018	Water	J3169-01	pH	Cool 4 deg C	WOOD06	N13	GMW-7	05/25/2018	SM 4500-PH B
05/25/2018	Water	J3169-04	pH	Cool 4 deg C	WOOD06	N13	GMW-X	05/25/2018	SM 4500-PH B
05/25/2018	Water	J3169-07	pH	Cool 4 deg C	WOOD06	N13	GMW-11	05/25/2018	SM 4500-PH B
05/25/2018	Water	J3169-10	pH	Cool 4 deg C	WOOD06	N13	FIELD-BLANK	05/25/2018	SM 4500-PH B
05/25/2018	Water	J3169-14	pH	Cool 4 deg C	WOOD06	N13	GMW-1	05/25/2018	SM 4500-PH B

Date/Time 05/27/18 8:40 AM

Received by: SP

Relinquished by: SP

Date/Time 05/27/18 9:15 AM

Received by: SP

Relinquished by: SP