


7470A

LB95668
INSTRUMENT ID: CV 4

Linear

 μ Abs.:

39390



A= 0.0000e+000

B= 2.5308e-004

C= 5.4064e-004

Rho= 0.9999849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-76	0.000	-76				
0.2	0.200	0.205	0.005	806	0.0 %	806				
2.5	2.500	2.498	-0.002	9867	0.0 %	9867				
5.0	5.000	5.018	0.018	19825	0.0 %	19825				
7.5	7.500	7.529	0.029	29747	0.0 %	29747				
10.0	10.000	9.970	-0.030	39390	0.0 %	39390				

LB95668 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 May 2018 09:25:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-76	-	0.0000	7470A	PPB	29 May 2018 09:33:27	S	Std
0.2 - 1	50.2	806	-	0.2000	7470A	PPB	29 May 2018 09:35:35	S	Std
2.5 - 1	50.5	9867	-	2.5000	7470A	PPB	29 May 2018 09:37:43	S	Std
5.0 - 1	55	19825	-	5.0000	7470A	PPB	29 May 2018 09:39:51	S	Std
7.5 - 1	57.5	29747	-	7.5000	7470A	PPB	29 May 2018 09:41:59	S	Std
10.0 - 1	510	39390	-	10.0000	7470A	PPB	29 May 2018 09:49:01	S	Std
ICV70 - 1	ICV70	14972	3.7897	-7470A	PPB	PPB	29 May 2018 09:51:34	U	SMPL
ICB70 - 1	ICB70	-294	-0.0739	-7470A	PPB	PPB	29 May 2018 09:53:40	U	SMPL
CCV76 - 1	CCV76	20115	5.0913	-7470A	PPB	PPB	29 May 2018 09:55:48	U	SMPL
CCB76 - 1	CCB76	-250	-0.0627	-7470A	PPB	PPB	29 May 2018 09:57:55	U	SMPL
CRA - 1	CRA	806	0.2045	-7470A	PPB	PPB	29 May 2018 10:00:02	U	SMPL
HighStd - 1	HighStd	39508	9.9994	-7470A	PPB	PPB	29 May 2018 10:02:08	U	SMPL
ChkStd - 1	ChkStd	28010	7.0894	-7470A	PPB	PPB	29 May 2018 10:04:15	U	SMPL
PB109714BL - 1	PBW	-165	-0.0412	-7470A	PPB	PPB	29 May 2018 10:08:58	U	SMPL
PB109714BS - 1	LCSW	14980	3.7917	-7470A	PPB	PPB	29 May 2018 10:11:05	U	SMPL
PB109714TB - 1	PB109660TB	-223	-0.0559	-7470A	PPB	PPB	29 May 2018 10:13:13	U	SMPL
J2679-26 - 1	EO-03-052418-A	-96	-0.0238	-7470A	PPB	PPB	29 May 2018 10:15:21	U	SMPL
J2679-26DUP - 1	EO-03-052418-ADUP	-62	-0.0152	-7470A	PPB	PPB	29 May 2018 10:17:30	U	SMPL
J2679-26MS - 1	EO-03-052418-AMS	14923	3.7773	-7470A	PPB	PPB	29 May 2018 10:19:38	U	SMPL
J2679-26MSD - 1	EO-03-052418-AMSD	15222	3.8530	-7470A	PPB	PPB	29 May 2018 10:21:45	U	SMPL
CCV77 - 1	CCV77	20440	5.1736	-7470A	PPB	PPB	29 May 2018 10:23:51	U	SMPL
CCB77 - 1	CCB77	-280	-0.0703	-7470A	PPB	PPB	29 May 2018 10:25:58	U	SMPL
J2679-28 - 1	EO-03-052418-B	-131	-0.0326	-7470A	PPB	PPB	29 May 2018 10:28:06	U	SMPL
J2679-30 - 1	EO-03-052418-C	-105	-0.0260	-7470A	PPB	PPB	29 May 2018 10:30:12	U	SMPL
J3114-03 - 1	NYS-RT-3460	-83	-0.0205	-7470A	PPB	PPB	29 May 2018 10:32:19	U	SMPL
J3114-07 - 1	NYS-RBR-200035	-104	-0.0258	-7470A	PPB	PPB	29 May 2018 10:34:26	U	SMPL
J3128-03 - 1	TJ-SOIL-RO-72-11995	-98	-0.0243	-7470A	PPB	PPB	29 May 2018 10:36:34	U	SMPL
J3128-06 - 1	TJ-OILY-GRAVEL	-79	-0.0195	-7470A	PPB	PPB	29 May 2018 10:38:41	U	SMPL
J3130-02 - 1	TP-5-WC-1	-73	-0.0179	-7470A	PPB	PPB	29 May 2018 10:40:49	U	SMPL
J3130-04 - 1	TP-5-WC-2	-59	-0.0144	-7470A	PPB	PPB	29 May 2018 10:42:57	U	SMPL
J3130-06 - 1	TP-5-WC-3	-77	-0.0189	-7470A	PPB	PPB	29 May 2018 10:45:06	U	SMPL
J3130-08 - 1	TP-5-WC-4	-57	-0.0139	-7470A	PPB	PPB	29 May 2018 10:47:13	U	SMPL
CCV78 - 1	CCV78	19957	5.0513	-7470A	PPB	PPB	29 May 2018 10:49:20	U	SMPL
CCB78 - 1	CCB78	-246	-0.0617	-7470A	PPB	PPB	29 May 2018 10:51:27	U	SMPL
PB109715BL - 1	PBW	-105	-0.0260	-7470A	PPB	PPB	29 May 2018 10:53:35	U	SMPL
PB109715BS - 1	LCSW	15425	3.9044	-7470A	PPB	PPB	29 May 2018 10:55:42	U	SMPL
PB109715TB - 1	PB109666TB	-161	-0.0402	-7470A	PPB	PPB	29 May 2018 10:57:49	U	SMPL
J3116-02 - 1	ETS	-63	-0.0154	-7470A	PPB	PPB	29 May 2018 10:59:57	U	SMPL
J3116-02DUP - 1	ETSDUP	-50	-0.0121	-7470A	PPB	PPB	29 May 2018 11:02:04	U	SMPL
J3116-02MS - 1	ETSMS	14895	3.7702	-7470A	PPB	PPB	29 May 2018 11:04:12	U	SMPL
J3116-02MSD - 1	ETSMSD	14539	3.6801	-7470A	PPB	PPB	29 May 2018 11:06:19	U	SMPL
J3125-02 - 1	72-12013	-133	-0.0331	-7470A	PPB	PPB	29 May 2018 11:08:28	U	SMPL
PB109716BL - 1	PBW	-57	-0.0139	-7470A	PPB	PPB	29 May 2018 11:10:37	U	SMPL
PB109716BS - 1	LCSW	15504	3.9243	-7470A	PPB	PPB	29 May 2018 11:12:44	U	SMPL
CCV79 - 1	CCV79	19745	4.9977	-7470A	PPB	PPB	29 May 2018 11:14:51	U	SMPL
CCB79 - 1	CCB79	-302	-0.0759	-7470A	PPB	PPB	29 May 2018 11:16:58	U	SMPL
J3139-03 - 1	BS-BASELINEWATER	137	0.0352	-7470A	PPB	PPB	29 May 2018 11:19:06	U	SMPL
J3139-03DUP - 1	BS-BASELINEWATERDUP	124	0.0319	-7470A	PPB	PPB	29 May 2018 11:21:13	U	SMPL
J3139-03MS - 1	BS-BASELINEWATERMS	15546	3.9350	-7470A	PPB	PPB	29 May 2018 11:23:20	U	SMPL
J3139-03MSD - 1	BS-BASELINEWATERMSD	15728	3.9810	-7470A	PPB	PPB	29 May 2018 11:25:27	U	SMPL
J2679-26LX5 - 1		-258	-0.0648	-7470A	PPB	PPB	29 May 2018 11:27:36	U	SMPL
J2679-26A - 1		14314	3.6232	-7470A	PPB	PPB	29 May 2018 11:29:45	U	SMPL
J3116-02LX5 - 1		-221	-0.0554	-7470A	PPB	PPB	29 May 2018 11:31:54	U	SMPL
J3116-02A - 1		14628	3.7026	-7470A	PPB	PPB	29 May 2018 11:34:02	U	SMPL
J3139-03LX5 - 1		-211	-0.0529	-7470A	PPB	PPB	29 May 2018 11:36:10	U	SMPL
J3139-03A - 1		14116	3.5731	-7470A	PPB	PPB	29 May 2018 11:38:17	U	SMPL
CCV80 - 1	CCV80	19480	4.9306	-7470A	PPB	PPB	29 May 2018 11:40:24	U	SMPL
CCB80 - 1	CCB80	-330	-0.0830	-7470A	PPB	PPB	29 May 2018 11:42:33	U	SMPL
J3078-15 - 1	PILE-2	-143	-0.0357	-7470A	PPB	PPB	29 May 2018 12:03:57	U	SMPL
CCV81 - 1	CCV81	19896	5.0359	-7470A	PPB	PPB	29 May 2018 12:06:03	U	SMPL
CCB81 - 1	CCB81	-262	-0.0658	-7470A	PPB	PPB	29 May 2018 12:08:11	U	SMPL