

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB95675

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP61786
calibration std. phosphate 1 ppm	WP64792
calibration std. phosphate 0.5 ppm	WP64791
calibration std. phosphate 0.3 ppm	WP64790
calibration std. 0 ppm	WP64876
phosphate ICV std.	WP64794
phosphate CCV std.	WP64793
calibration std. phosphate 0.1 ppm	WP64789
calibration std. phosphate 0.05 ppm	WP64788
Combined reagent	WP64875
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0011Slope: 0.6511Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	05/29/2018	12:07
2	CAL2	0.05	1	50	50	0.033	05/29/2018	12:07
3	CAL3	0.10	1	50	50	0.065	05/29/2018	12:08
4	CAL4	0.30	1	50	50	0.186	05/29/2018	12:09
5	CAL5	0.50	1	50	50	0.329	05/29/2018	12:09
6	CAL6	1.00	1	50	50	0.650	05/29/2018	12:10



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB95675

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.496	05/29/2018	12:10
2	ICB		1	50	50	0.000	0.002	05/29/2018	12:11
3	CCV1	0.50	1	50	50	0.326	0.502	05/29/2018	12:12
4	CCB1		1	50	50	0.000	0.002	05/29/2018	12:12
5	PB109764BL		1	50	50	0.000	0.002	05/29/2018	12:13
6	PB109764BS	0.50	1	50	50	0.323	0.498	05/29/2018	12:13
7	J3122-01		1	50	50	1.206	1.854	05/29/2018	12:14
8	J3122-01DUP		1	50	50	1.144	1.759	05/29/2018	12:14
9	J3122-01MS	0.50	1	50	50	1.639	2.519	05/29/2018	12:15
10	J3122-01MSD	0.50	1	50	50	1.630	2.505	05/29/2018	12:16
11	J3122-01		10	50	50	0.402	0.619	05/29/2018	12:30
12	J3122-01DUP		10	50	50	0.393	0.605	05/29/2018	12:31
13	J3122-01MS	0.50	10	50	50	0.432	0.665	05/29/2018	12:31
14	J3122-01MSD	0.50	10	50	50	0.428	0.659	05/29/2018	12:32
15	CCV2	0.50	1	50	50	0.327	0.504	05/29/2018	12:33
16	CCB2		1	50	50	0.000	0.002	05/29/2018	12:33