



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95686

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	05/29/2018	09:15
2	J3139-04		1	>100.0	>212.0	>212.0	05/29/2018	09:40
3	J3139-04DUP		1	>100.0	>212.0	>212.0	05/29/2018	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95686

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-052918

WorkList ID : 112682

Date : 5/29/2018 8:16:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2018	Water	J3139-04 C	Flash Point	Cool 4 deg C	PSEG01	O23	BS-BASELINEWATER	05/25/2018	1010A

Date/Time 05/29/18 8:30 AM

Received by: KS

Relinquished by: JP

Date/Time 05/29/18 10:15 AM

Received by: JP

Relinquished by: KS