



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 05/30/2018TEMP1 IN: 104 05/30/2018 07:30 TEMP OUT: 104 05/30/2018 08:32TEMP2 IN: 104 05/30/2018 09:00 TEMP OUT: 104 05/30/2018 10:05TEMP3 IN: 105 05/30/2018 11:15 TEMP OUT: 105 05/30/2018 12:30TEMP4 IN: 104 05/30/2018 13:45 TEMP OUT: 104 05/30/2018 14:33Run Number: LB95696BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95696BL	LB95696BL	1.3658	1.3658	100	1.3658	1.3658	1.3658	0.0000	0
2	LB95696BS	LB95696BS	1.4355	1.4355	100	1.4902	1.4902	1.4902	0.0547	547
3	J3192-01	14B-1	1.4075	1.4075	20	1.4589	1.4589	1.4589	0.0514	2570
4	J3192-01DUP	14B-1DUP	1.4137	1.4137	20	1.4664	1.4664	1.4664	0.0527	2635
5	J3192-02	14B-2	1.3878	1.3878	20	1.4390	1.4390	1.4390	0.0512	2560
6	J3192-03	14B-3	1.3919	1.3919	20	1.4449	1.4449	1.4449	0.0530	2650
7	J3192-04	14B-4	1.4030	1.4030	20	1.4776	1.4776	1.4776	0.0746	3730
8	J3194-01	001M-WILLETS-PT-BLVD (APR)	1.3979	1.3979	200	1.4057	1.4057	1.4057	0.0078	39
9	J3194-02	001M-WILLETS-PT-BLVD (MAY)	1.4149	1.4149	200	1.4211	1.4211	1.4211	0.0062	31

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

LB 95696

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J3199

WorkList ID : 112719

Date : 5/30/2018 7:36:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2018	Water	J3192-01-10	TSS	Cool 4 deg C	NEWY17	N62	14B-1	05/29/2018	SM2540 D
06/05/2018	Water	J3192-02-10	TSS	Cool 4 deg C	NEWY17	N62	14B-2	05/29/2018	SM2540 D
06/05/2018	Water	J3192-03-10	TSS	Cool 4 deg C	NEWY17	N62	14B-3	05/29/2018	SM2540 D
06/05/2018	Water	J3192-04-10	TSS	Cool 4 deg C	NEWY17	N62	14B-4	05/29/2018	SM2540 D
05/30/2018	Water	J3194-01-10	TSS	Cool 4 deg C	TULL01	N52	001M-WILLETS-PT-BLVD(API	05/24/2018	SM2540 D
05/31/2018	Water	J3194-02-10	TSS	Cool 4 deg C	TULL01	N52	001M-WILLETS-PT-BLVD(MA	05/24/2018	SM2540 D

Date/Time 05/30/18 7:40 AM
 Received by: JS
 Relinquished by: JS

Date/Time 05/30/18 10:30 AM
 Received by: JS
 Relinquished by: JS