



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95715

Slope : 97.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/30/2018	12:10
2	CAL2	1	Water			21.3	7.00	05/30/2018	12:12
3	CAL3	1	Water			21.4	10.03	05/30/2018	12:14
4	ICV	1	Water			21.5	7.00	05/30/2018	12:16
5	CCV1	1	Water			21.6	2.00	05/30/2018	12:18
6	J3204-01	1	Water			22.3	4.31	05/30/2018	12:20
7	J3204-01DUP	1	Water			22.4	4.32	05/30/2018	12:22
8	CCV2	1	Water			21.5	12.00	05/30/2018	12:24

WORKLIST(Hardcopy Internal Chain)

LB95715

WorkList Name : pH j3204

WorkList ID : 112755

Date : 5/30/2018 11:54:04 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2018	Water	J3204-01	pH	Cool 4 deg C	STER12	N13	CELL-POOL-WATER-2NDQ	05/29/2018	SM 4500-PH B

Date/Time 05/30/18 12:05 PM

Received by: SP

Relinquished by: CP

Date/Time 05/30/18 12:35 PM

Received by: CP

Relinquished by: SP