


7470A

LB 95723

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.5876e-004

C= 9.3379e-004

Rho= 0.9999815

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-51	0.000	-51				
0.2	0.200	0.208	0.008	800	0.0 %	800				
2.5	2.500	2.533	0.033	9786	0.0 %	9786				
5.0	5.000	4.961	-0.039	19167	0.0 %	19167				
7.5	7.500	7.498	-0.002	28975	0.0 %	28975				
10.0	10.000	10.012	0.012	38689	0.0 %	38689				

LB95723 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 May 2018 09:51:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-51	-	0.0000	7470A	PPB	31 May 2018 10:05:36	S	Std
0.2 - 1	50.2	800	-	0.2000	7470A	PPB	31 May 2018 10:07:44	S	Std
2.5 - 1	52.5	9786	-	2.5000	7470A	PPB	31 May 2018 10:09:52	S	Std
5.0 - 1	55	19167	-	5.0000	7470A	PPB	31 May 2018 10:12:00	S	Std
7.5 - 1	57.5	28975	-	7.5000	7470A	PPB	31 May 2018 10:14:09	S	Std
10.0 - 1	510	38689	-	10.0000	7470A	PPB	31 May 2018 10:18:43	S	Std
ICV74 - 1	ICV74	14337	3.7108	-	7470A	PPB	31 May 2018 10:21:12	U	SMPL
ICB74 - 1	ICB74	-215	-0.0547	-	7470A	PPB	31 May 2018 10:23:19	U	SMPL
CCV05 - 1	CCV05	19349	5.0077	-	7470A	PPB	31 May 2018 10:25:26	U	SMPL
CCB05 - 1	CCB05	-187	-0.0475	-	7470A	PPB	31 May 2018 10:27:32	U	SMPL
CRA - 1	CRA	817	0.2123	-	7470A	PPB	31 May 2018 10:29:40	U	SMPL
HighStd - 1	HighStd	38431	9.9453	-	7470A	PPB	31 May 2018 10:31:46	U	SMPL
ChkStd - 1	ChkStd	26130	6.7623	-	7470A	PPB	31 May 2018 10:33:53	U	SMPL
PB109805BL - 1	PBW	-255	-0.0650	-	7470A	PPB	31 May 2018 10:36:03	U	SMPL
PB109805BS - 1	LCSW	13478	3.4885	-	7470A	PPB	31 May 2018 10:38:12	U	SMPL
PB109805TB - 1	PB109765TB	-101	-0.0252	-	7470A	PPB	31 May 2018 10:40:19	U	SMPL
J3162-06 - 1	S-06	-29	-0.0066	-	7470A	PPB	31 May 2018 10:42:28	U	SMPL
J3162-06DUP - 1	S-06DUP	-48	-0.0115	-	7470A	PPB	31 May 2018 10:44:36	U	SMPL
J3162-06MS - 1	S-06MS	14143	3.6606	-	7470A	PPB	31 May 2018 10:46:45	U	SMPL
J3162-06MSD - 1	S-06MSD	13189	3.4137	-	7470A	PPB	31 May 2018 10:48:52	U	SMPL
CCV06 - 1	CCV06	19552	5.0602	-	7470A	PPB	31 May 2018 10:50:58	U	SMPL
CCB06 - 1	CCB06	-169	-0.0428	-	7470A	PPB	31 May 2018 10:53:05	U	SMPL
J3186-06 - 1	S-6-COMP	-3	0.0002	-	7470A	PPB	31 May 2018 10:55:13	U	SMPL
J3188-02 - 1	ES-CONCRETE	-23	-0.0050	-	7470A	PPB	31 May 2018 10:57:20	U	SMPL
J3190-01 - 1	RED-INSIDE-WALL	-4	-0.0001	-	7470A	PPB	31 May 2018 10:59:26	U	SMPL
J3190-02 - 1	TAN-OUTSIDE-WALL	-61	-0.0149	-	7470A	PPB	31 May 2018 11:01:33	U	SMPL
J3191-03 - 1	BRICKWALL-DRUM	-25	-0.0055	-	7470A	PPB	31 May 2018 11:03:41	U	SMPL
J3197-04 - 1	TP-2-S	290	0.0760	-	7470A	PPB	31 May 2018 11:05:48	U	SMPL
J3197-08 - 1	TP-2-D	6	0.0025	-	7470A	PPB	31 May 2018 11:07:56	U	SMPL
J3162-06LX5 - 1		-41	-0.0097	-	7470A	PPB	31 May 2018 11:10:04	U	SMPL
J3162-06A - 1		14269	3.6932	-	7470A	PPB	31 May 2018 11:12:12	U	SMPL
CCV07 - 1	CCV07	19134	4.9520	-	7470A	PPB	31 May 2018 11:14:19	U	SMPL
CCB07 - 1	CCB07	-204	-0.0519	-	7470A	PPB	31 May 2018 11:16:26	U	SMPL