


7471B

LB95731
INSTRUMENT ID: CV2

Linear

 μ Abs.:

42741



A= 0.0000e+000

B= 2.3757e-004

C= -1.8411e-002

Rho= 0.9993481

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.016	-0.016	12	0.000	12				
0.2	0.200	0.183	-0.017	846	0.0 %	846				
2.5	2.500	2.609	0.109	11060	0.0 %	11060				
5.0	5.000	5.057	0.057	21366	0.0 %	21366				
7.5	7.500	7.231	-0.269	30515	0.0 %	30515				
10.0	10.000	10.135	0.135	42741	0.0 %	42741				

LB95731 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 May 2018 11:47:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1		12	-	0.0000	7471B	PPB	31 May 2018 11:48:56	S	Std
0.2 - 1		846	-	0.2000	7471B	PPB	31 May 2018 11:51:13	S	Std
2.5 - 1		11060	-	2.5000	7471B	PPB	31 May 2018 11:53:31	S	Std
5.0 - 1		21366	-	5.0000	7471B	PPB	31 May 2018 11:55:48	S	Std
7.5 - 1		30515	-	7.5000	7471B	PPB	31 May 2018 11:58:05	S	Std
10.0 - 1		42741	-	10.0000	7471B	PPB	31 May 2018 12:00:22	S	Std
ICV75 - 1	ICV75	16636	3.9338	-	7471B	PPB	31 May 2018 12:11:02	U	SMPL
ICB75 - 1	ICB75	-134	-0.0502	-	7471B	PPB	31 May 2018 12:13:18	U	SMPL
CCV08 - 1	CCV08	21321	5.0468	-	7471B	PPB	31 May 2018 12:15:33	U	SMPL
CCB08 - 1	CCB08	-177	-0.0605	-	7471B	PPB	31 May 2018 12:17:49	U	SMPL
CRA - 1	CRA	826	0.1778	-	7471B	PPB	31 May 2018 12:20:05	U	SMPL
HighStd - 1	HighStd	43600	10.3395	-	7471B	PPB	31 May 2018 12:22:21	U	SMPL
ChkStd - 1	ChkStd	30289	7.1773	-	7471B	PPB	31 May 2018 12:24:37	U	SMPL
PB109812BL - 1	PBS	-254	-0.0788	-	7471B	PPB	31 May 2018 12:26:56	U	SMPL
PB109812BS - 1	LCSS	17078	4.0388	-	7471B	PPB	31 May 2018 12:29:14	U	SMPL
J2669-11 - 1	SU-03-053018	1610	0.3641	-	7471B	PPB	31 May 2018 12:31:31	U	SMPL
J2678-05 - 1	OK-01-053018	1908	0.4349	-	7471B	PPB	31 May 2018 12:33:48	U	SMPL
J3015-04 - 1	DUP2-051718	593	0.1225	-	7471B	PPB	31 May 2018 12:36:06	U	SMPL
J3188-01 - 1	ES-CONCRETE	465	0.0921	-	7471B	PPB	31 May 2018 12:38:23	U	SMPL
J3191-02 - 1	BRICKWALL-DRUM	996	0.2182	-	7471B	PPB	31 May 2018 12:40:39	U	SMPL
CCV09 - 1	CCV09	21414	5.0689	-	7471B	PPB	31 May 2018 12:42:56	U	SMPL
CCB09 - 1	CCB09	-160	-0.0564	-	7471B	PPB	31 May 2018 12:45:11	U	SMPL
J3197-03 - 1	TP-2-S	2923	0.6760	-	7471B	PPB	31 May 2018 12:47:27	U	SMPL
J3197-07 - 1	TP-2-D	472	0.0937	-	7471B	PPB	31 May 2018 12:49:43	U	SMPL
J3202-01 - 1	SP	3360	0.7798	-	7471B	PPB	31 May 2018 12:51:59	U	SMPL
J3202-01DUP - 1	SPDUP	3723	0.8661	-	7471B	PPB	31 May 2018 12:54:16	U	SMPL
J3202-01MS - 1	SPMS	20195	4.7793	-	7471B	PPB	31 May 2018 12:56:32	U	SMPL
J3202-01MSD - 1	SPMSD	20229	4.7873	-	7471B	PPB	31 May 2018 12:58:50	U	SMPL
J3203-01 - 1	TP-2-WC-1	323	0.0583	-	7471B	PPB	31 May 2018 13:01:07	U	SMPL
J3203-04 - 1	TP-4-WC-1	474	0.0942	-	7471B	PPB	31 May 2018 13:03:25	U	SMPL
J3217-01 - 1	SP-01	622	0.1294	-	7471B	PPB	31 May 2018 13:05:43	U	SMPL
J3217-05 - 1	DRUM-1	-14	-0.0217	-	7471B	PPB	31 May 2018 13:07:59	U	SMPL
CCV10 - 1	CCV10	21565	5.1047	-	7471B	PPB	31 May 2018 13:10:15	U	SMPL
CCB10 - 1	CCB10	-90	-0.0398	-	7471B	PPB	31 May 2018 13:12:32	U	SMPL
J3217-08 - 1	SP-MANHOLE-TRENCH-1	600	0.1241	-	7471B	PPB	31 May 2018 13:14:48	U	SMPL
J3217-12 - 1	SP-MANHOLE-TRENCH-2	584	0.1203	-	7471B	PPB	31 May 2018 13:17:04	U	SMPL
J3218-01 - 1	OUTSIDE-WELL-SP	2262	0.5190	-	7471B	PPB	31 May 2018 13:19:21	U	SMPL
J3218-05 - 1	INSIDE-WELL-SP	2968	0.6867	-	7471B	PPB	31 May 2018 13:21:37	U	SMPL
J3219-03 - 1	TP-3	49959	11.8502	-	7471B	PPB	31 May 2018 13:23:54	U	SMPL
J3220-01 - 1	EP-14	257	0.0426	-	7471B	PPB	31 May 2018 13:26:11	U	SMPL
J3220-02 - 1	EP-22	464	0.0918	-	7471B	PPB	31 May 2018 13:28:30	U	SMPL
J3202-01LX5 - 1		600	0.1241	-	7471B	PPB	31 May 2018 13:30:47	U	SMPL
J3202-01A - 1		21718	5.1411	-	7471B	PPB	31 May 2018 13:33:05	U	SMPL
CCV11 - 1	CCV11	21442	5.0755	-	7471B	PPB	31 May 2018 13:35:21	U	SMPL
CCB11 - 1	CCB11	-154	-0.0550	-	7471B	PPB	31 May 2018 13:37:38	U	SMPL
J3219-03DLX2 - 1		23950	5.6713	-	7471B	PPB	31 May 2018 13:45:45	U	SMPL
CCV12 - 1	CCV12	21467	5.0815	-	7471B	PPB	31 May 2018 13:48:01	U	SMPL
CCB12 - 1	CCB12	-140	-0.0517	-	7471B	PPB	31 May 2018 13:50:18	U	SMPL