


245.1

LB95741
INSTRUMENT ID: CV2

Linear

 μ Abs.:

33193



A= 0.0000e+000

B= 2.9862e-004

C= -4.1755e-004

Rho= 0.9998863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-65	0.000	-65				
0.05	0.050	0.061	0.011	206	0.0 %	206				
0.2	0.200	0.192	-0.008	646	0.0 %	646				
2.5	2.500	2.478	-0.022	8301	0.0 %	8301				
5.0	5.000	5.012	0.012	16785	0.0 %	16785				
7.5	7.500	7.614	0.114	25498	0.0 %	25498				
10.0	10.000	9.912	-0.088	33193	0.0 %	33193				

LB95741 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 May 2018 15:34:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-65	-	0.0000	245.1	PPB	31 May 2018 15:41:00	S	Std
0.05 - 1	50.05	206	-	0.0500	245.1	PPB	31 May 2018 15:43:08	S	Std
0.2 - 1	50.2	646	-	0.2000	245.1	PPB	31 May 2018 15:45:18	S	Std
2.5 - 1	52.5	8301	-	2.5000	245.1	PPB	31 May 2018 15:47:26	S	Std
5.0 - 1	55	16785	-	5.0000	245.1	PPB	31 May 2018 15:49:34	S	Std
7.5 - 1	57.5	25498	-	7.5000	245.1	PPB	31 May 2018 15:51:41	S	Std
10.0 - 1	510	33193	-	10.0000	245.1	PPB	31 May 2018 16:00:55	S	Std
ICV76 - 1	ICV76	12979	3.8754	-245.1		PPB	31 May 2018 16:04:02	U	SMPL
ICB76 - 1	ICB76	-230	-0.0691	-245.1		PPB	31 May 2018 16:06:08	U	SMPL
CCV13 - 1	CCV13	17083	5.1010	-245.1		PPB	31 May 2018 16:14:05	U	SMPL
CCB13 - 1	CCB13	-268	-0.0804	-245.1		PPB	31 May 2018 16:16:11	U	SMPL
CRA - 1	CRA	138	0.0408	-245.1		PPB	31 May 2018 16:20:58	U	SMPL
HighStd - 1	HighStd	33190	9.9109	-245.1		PPB	31 May 2018 16:23:04	U	SMPL
ChkStd - 1	ChkStd	23048	6.8823	-245.1		PPB	31 May 2018 16:25:11	U	SMPL
PB109806BL - 1	PB109806BL	-221	-0.0664	-245.1		PPB	31 May 2018 16:27:21	U	SMPL
PB109806BS - 1	PB109806BS	13582	4.0555	-245.1		PPB	31 May 2018 16:29:29	U	SMPL
J3210-01 - 1	052918-DBRI-F-D-B	-185	-0.0557	-245.1		PPB	31 May 2018 16:31:36	U	SMPL
J3210-02 - 1	052918-DBRI-F-D-M	-84	-0.0255	-245.1		PPB	31 May 2018 16:33:45	U	SMPL
J3210-03 - 1	052918-DBRI-F-D-S	-104	-0.0315	-245.1		PPB	31 May 2018 16:35:53	U	SMPL
J3210-04 - 1	052918-DBRI-F-U-B	-77	-0.0234	-245.1		PPB	31 May 2018 16:38:02	U	SMPL
J3210-05 - 1	052918-DBRI-F-U-M	-85	-0.0258	-245.1		PPB	31 May 2018 16:40:08	U	SMPL
CCV14 - 1	CCV14	16770	5.0075	-245.1		PPB	31 May 2018 16:42:15	U	SMPL
CCB14 - 1	CCB14	-236	-0.0709	-245.1		PPB	31 May 2018 16:44:22	U	SMPL
J3210-06 - 1	052918-DBRI-F-U-S	-109	-0.0330	-245.1		PPB	31 May 2018 16:46:30	U	SMPL
J3210-07 - 1	052918-JETT-E-D-B	-112	-0.0339	-245.1		PPB	31 May 2018 16:48:37	U	SMPL
J3210-08 - 1	052918-JETT-E-D-M	-97	-0.0294	-245.1		PPB	31 May 2018 16:50:43	U	SMPL
J3210-09 - 1	052918-JETT-E-D-S	-76	-0.0231	-245.1		PPB	31 May 2018 16:52:50	U	SMPL
J3210-10 - 1	052918-JETT-E-U-B	-84	-0.0255	-245.1		PPB	31 May 2018 16:54:58	U	SMPL
J3210-11 - 1	052918-JETT-E-U-M	-78	-0.0237	-245.1		PPB	31 May 2018 16:57:05	U	SMPL
J3210-12 - 1	052918-JETT-E-U-S	-116	-0.0351	-245.1		PPB	31 May 2018 16:59:13	U	SMPL
J3210-12DUP - 1	052918-JETT-E-U-SDUP	-90	-0.0273	-245.1		PPB	31 May 2018 17:01:21	U	SMPL
J3210-12MS - 1	052918-JETT-E-U-SMS	16050	4.7925	-245.1		PPB	31 May 2018 17:03:29	U	SMPL
J3210-12MSD - 1	052918-JETT-E-U-SMSD	15172	4.5303	-245.1		PPB	31 May 2018 17:05:36	U	SMPL
CCV15 - 1	CCV15	16847	5.0305	-245.1		PPB	31 May 2018 17:07:43	U	SMPL
CCB15 - 1	CCB15	-266	-0.0799	-245.1		PPB	31 May 2018 17:09:51	U	SMPL
J3212-12 - 1	14B(1-4)-COMP	296	0.0880	-245.1		PPB	31 May 2018 17:11:59	U	SMPL
J3215-01 - 1	COMP	-43	-0.0133	-245.1		PPB	31 May 2018 17:14:06	U	SMPL
J3210-12LX5 - 1		-84	-0.0255	-245.1		PPB	31 May 2018 17:16:13	U	SMPL
J3210-12A - 1		16001	4.7779	-245.1		PPB	31 May 2018 17:18:19	U	SMPL
CCV16 - 1	CCV16	17020	5.0822	-245.1		PPB	31 May 2018 17:20:26	U	SMPL
CCB16 - 1	CCB16	-234	-0.0703	-245.1		PPB	31 May 2018 17:22:34	U	SMPL
J3171-10 - 1	052418-JETT-F-U-B	-60	-0.0183	-245.1		PPB	31 May 2018 18:14:12	U	SMPL
CCV17 - 1	CCV17	16663	4.9756	-245.1		PPB	31 May 2018 18:16:20	U	SMPL
CCB17 - 1	CCB17	-184	-0.0554	-245.1		PPB	31 May 2018 18:18:28	U	SMPL