



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95747

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/31/2018	16:40
2	CAL2	1	Water			21.4	7.00	05/31/2018	16:42
3	CAL3	1	Water			21.6	10.03	05/31/2018	16:44
4	ICV	1	Water			21.3	7.00	05/31/2018	16:46
5	CCV1	1	Water			21.5	2.00	05/31/2018	17:24
6	J2969-01	1	Water			23.5	4.67	05/31/2018	17:26
7	J2969-01DUP	1	Water			23.6	4.68	05/31/2018	17:28
8	J3237-04	1	Water			24.3	7.27	05/31/2018	17:30
9	CCV2	1	Water			21.6	12.00	05/31/2018	17:34

LB95747

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity053118-1

WorkList ID : 112821

Date : 5/31/2018 3:40:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/16/2018	Water	J2969-01	Corrosivity	Cool 4 deg C	WARR01	J43	UNKNOWN-1	05/16/2018	9040C
05/31/2018	Water	J3237-04	Corrosivity	Cool 4 deg C	PSEG01	I43	AG-BASELINEWATER	05/31/2018	9040C

Date/Time 05/31/18 4:30 PM

Received by: SP

Relinquished by: CP

Date/Time 05/31/18 6:00 PM

Received by: CP

Relinquished by: SP