



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95748

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/31/2018	16:40
2	CAL2	1	Water			21.4	7.00	05/31/2018	16:42
3	CAL3	1	Water			21.6	10.03	05/31/2018	16:46
4	ICV	1	Water			21.3	7.00	05/31/2018	16:46
5	CCV1	1	Water			21.5	2.00	05/31/2018	16:48
6	J3228-01	1	Solid	20.02	20	23.3	6.07	05/31/2018	17:35
7	J3228-01DUP	1	Solid	20.04	20	23.3	6.08	05/31/2018	17:37
8	CCV2	1	Water			21.4	12.00	05/31/2018	17:39

WORKLIST(Hardcopy Internal Chain)

LB 95748

WorkList Name : pH053118

WorkList ID : 112822

Date : 5/31/2018 3:40:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/30/2018	Solid	J3228-01	pH	Cool 4 deg C	TULL02	N41	FK-PS-01-028	05/30/2018	9045C

Date/Time 05/31/18 4:30 PM
 Received by: SP
 Relinquished by: CP

Date/Time 05/31/18 6:00 PM
 Received by: CP
 Relinquished by: SP