



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95749

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	05/31/2018	16:40
2	CAL2	1	Water			21.4	7.00	05/31/2018	16:42
3	CAL3	1	Water			21.6	10.03	05/31/2018	16:44
4	ICV	1	Water			21.3	7.00	05/31/2018	16:46
5	CCV1	1	Water			21.5	2.00	05/31/2018	17:24
6	J3237-03	1	Water			24.3	7.27	05/31/2018	17:30
7	J3237-03DUP	1	Water			24.4	7.28	05/31/2018	17:32
8	CCV2	1	Water			21.5	12.00	05/31/2018	17:34

WORKLIST(Hardcopy Internal Chain)

LB 95749

WorkList Name : pH053118-1

WorkList ID : 112823

Date : 5/31/2018 3:41:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/31/2018	Water	J3237-03	pH	Cool 4 deg C	PSEG01	I43	AG-BASELINEWATER	05/31/2018	9040C

Date/Time 05/31/18 4:30 PM
 Received by: SP
 Relinquished by: CAP

Date/Time 05/31/18 6:00 PM
 Received by: CP
 Relinquished by: SP