

LB95753


7470A

INSTRUMENT ID: CVD

Linear

μ Abs.:

37880



A= 0.0000e+000

B= 2.6452e-004

C= 2.1841e-002

Rho= 0.9999255

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-109	0.000	-109				
0.2	0.200	0.200	-0.000	672	0.0 %	672				
2.5	2.500	2.512	0.012	9415	0.0 %	9415				
5.0	5.000	5.044	0.044	18987	0.0 %	18987				
7.5	7.500	7.409	-0.091	27927	0.0 %	27927				
10.0	10.000	10.042	0.042	37880	0.0 %	37880				

LB95753 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jun 2018 08:14:28

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SB	-109	-	0.00007470A	PPB		01 Jun 2018 08:19:36	S	Std
0.2 - 1	SB 2	672	-	0.20007470A	PPB		01 Jun 2018 08:21:44	S	Std
2.5 - 1	SB 2.5	9415	-	2.50007470A	PPB		01 Jun 2018 08:23:52	S	Std
5.0 - 1	SB 5	18987	-	5.00007470A	PPB		01 Jun 2018 08:26:01	S	Std
7.5 - 1	SB 7.5	27927	-	7.50007470A	PPB		01 Jun 2018 08:28:09	S	Std
10.0 - 1	SB 10	37880	-	10.00007470A	PPB		01 Jun 2018 08:30:16	S	Std
ICV77 - 1	ICV77	14681	3.9052	-7470A	PPB		01 Jun 2018 08:33:00	U	SMPL
ICB77 - 1	ICB77	-290	-0.0549	-7470A	PPB		01 Jun 2018 08:35:07	U	SMPL
CCV18 - 1	CCV18	18857	5.0099	-7470A	PPB		01 Jun 2018 08:37:14	U	SMPL
CCB18 - 1	CCB18	-261	-0.0472	-7470A	PPB		01 Jun 2018 08:39:21	U	SMPL
CRA - 1	CRA	723	0.2131	-7470A	PPB		01 Jun 2018 08:41:28	U	SMPL
HighStd - 1	HighStd	38202	10.1270	-7470A	PPB		01 Jun 2018 08:43:35	U	SMPL
ChkStd - 1	ChkStd	26687	7.0810	-7470A	PPB		01 Jun 2018 08:45:41	U	SMPL
PB109840BL - 1	PBW	-345	-0.0694	-7470A	PPB		01 Jun 2018 08:47:50	U	SMPL
PB109840BS - 1	LCSW	14053	3.7391	-7470A	PPB		01 Jun 2018 08:49:58	U	SMPL
PB109840TB - 1	PB109808TB	-182	-0.0263	-7470A	PPB		01 Jun 2018 08:52:06	U	SMPL
J2669-12 - 1	SU-03-053018	-126	-0.0115	-7470A	PPB		01 Jun 2018 08:54:13	U	SMPL
J2669-12DUP - 1	SU-03-053018DUP	-138	-0.0147	-7470A	PPB		01 Jun 2018 08:56:22	U	SMPL
J2669-12MS - 1	SU-03-053018MS	13649	3.6323	-7470A	PPB		01 Jun 2018 08:58:30	U	SMPL
J2669-12MSD - 1	SU-03-053018MSD	13310	3.5426	-7470A	PPB		01 Jun 2018 09:00:37	U	SMPL
CCV19 - 1	CCV19	18962	5.0376	-7470A	PPB		01 Jun 2018 09:02:43	U	SMPL
CCB19 - 1	CCB19	-306	-0.0591	-7470A	PPB		01 Jun 2018 09:04:50	U	SMPL
J2678-06 - 1	OK-01-053018	-144	-0.0162	-7470A	PPB		01 Jun 2018 09:06:58	U	SMPL
J3202-04 - 1	SP	-112	-0.0078	-7470A	PPB		01 Jun 2018 09:09:05	U	SMPL
J3203-03 - 1	TP-2-WC-1	-93	-0.0028	-7470A	PPB		01 Jun 2018 09:11:11	U	SMPL
J3203-06 - 1	TP-4-WC-1	-63	0.0052	-7470A	PPB		01 Jun 2018 09:13:19	U	SMPL
J3217-04 - 1	SP-01	-86	-0.0009	-7470A	PPB		01 Jun 2018 09:15:26	U	SMPL
J3217-07 - 1	DRUM-1	-135	-0.0139	-7470A	PPB		01 Jun 2018 09:17:33	U	SMPL
J3217-11 - 1	SP-MANHOLE-TRENCH-1	-108	-0.0067	-7470A	PPB		01 Jun 2018 09:19:41	U	SMPL
J3217-15 - 1	SP-MANHOLE-TRENCH-2	-80	0.0007	-7470A	PPB		01 Jun 2018 09:21:49	U	SMPL
J3218-04 - 1	OUTSIDE-WELL-SP	-40	0.0113	-7470A	PPB		01 Jun 2018 09:23:58	U	SMPL
J3218-08 - 1	INSIDE-WELL-SP	-78	0.0012	-7470A	PPB		01 Jun 2018 09:26:04	U	SMPL
CCV20 - 1	CCV20	18697	4.9675	-7470A	PPB		01 Jun 2018 09:28:11	U	SMPL
CCB20 - 1	CCB20	-279	-0.0520	-7470A	PPB		01 Jun 2018 09:30:18	U	SMPL
J3219-04 - 1	TP-3	-105	-0.0059	-7470A	PPB		01 Jun 2018 09:32:25	U	SMPL
PB109851BL - 1	PBW	-117	-0.0091	-7470A	PPB		01 Jun 2018 09:34:32	U	SMPL
PB109851BS - 1	LCSW	13200	3.5135	-7470A	PPB		01 Jun 2018 09:36:40	U	SMPL
J2969-01 - 1	UNKNOWN-1	-59	0.0062	-7470A	PPB		01 Jun 2018 09:38:47	U	SMPL
J2669-12LX5 - 1		-146	-0.0168	-7470A	PPB		01 Jun 2018 09:40:54	U	SMPL
J2669-12A - 1		14620	3.8891	-7470A	PPB		01 Jun 2018 09:43:02	U	SMPL
CCV21 - 1	CCV21	18453	4.9030	-7470A	PPB		01 Jun 2018 09:45:10	U	SMPL
CCB21 - 1	CCB21	-262	-0.0475	-7470A	PPB		01 Jun 2018 09:47:18	U	SMPL