



## Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95760

Ambient Barometric Pressure (mmHg): 755

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2168     |

| Seq | LabID       | True Value °F | DL | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 27.0       | 80.6      | 80.9            | 06/01/2018 | 10:50     |
| 2   | J2969-01    |               | 1  | 45.0       | 113.0     | 113.3           | 06/01/2018 | 11:15     |
| 3   | J3237-04    |               | 1  | >100.0     | >212.0    | >212.0          | 06/01/2018 | 11:40     |
| 4   | J3237-04DUP |               | 1  | >100.0     | >212.0    | >212.0          | 06/01/2018 | 12:10     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB 95760

## WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-060118

WorkList ID : 112850

Date : 6/1/2018 8:12:35 AM

| Due Date   | Matrix | Sample     | Test        | Preservative | Customer | Storage Location | Customer Sample  | Collect Date | Method |
|------------|--------|------------|-------------|--------------|----------|------------------|------------------|--------------|--------|
| 05/17/2018 | Water  | J2969-01 ✓ | Flash Point | Cool 4 deg C | WARR01   | J43              | UNKNOWN-1        | 05/16/2018   | 1010A  |
| 06/05/2018 | Water  | J3237-04 C | Flash Point | Cool 4 deg C | PSEG01   | I43              | AG-BASELINEWATER | 05/31/2018   | 1010A  |

Date/Time

Received by:

Relinquished by:

06/01/18 8:30 AM

KS

CP

Date/Time

Received by:

Relinquished by:

06/01/18 12:15 PM

CP

KS