



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95761

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	82.3	06/01/2018	09:05
2	J2969-03		1	>100.0	>212.0	>212.0	06/01/2018	09:40
3	J2969-03DUP		1	>100.0	>212.0	>212.0	06/01/2018	10:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB 915761

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2-060118

WorkList ID : 112851

Date : 6/1/2018 8:13:22 AM

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2018	Solid	J2969-03	✓	Flash Point	Cool 4 deg C	WARR01	J43	UNKNOWN-2	05/16/2018	1010A

Date/Time 06/01/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time

Received by: CP

Relinquished by: KS