

7470A


LB95783

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38611



A= 0.0000e+000

B= 2.5973e-004

C= -5.9192e-002

Rho= 0.9999451

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	105	0.000	105				
0.2	0.200	0.167	-0.033	871	0.0 %	871				
2.5	2.500	2.575	0.075	10143	0.0 %	10143				
5.0	5.000	5.014	0.014	19534	0.0 %	19534				
7.5	7.500	7.506	0.006	29129	0.0 %	29129				
10.0	10.000	9.969	-0.031	38611	0.0 %	38611				

LB95783 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jun 2018 09:06:40

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	105	-	0.0000	7470A	PPB	04 Jun 2018 09:15:57	S	Std
0.2 - 1	50.2	871	-	0.2000	7470A	PPB	04 Jun 2018 09:18:05	S	Std
2.5 - 1	52.5	10143	-	2.5000	7470A	PPB	04 Jun 2018 09:20:13	S	Std
5.0 - 1	55	19534	-	5.0000	7470A	PPB	04 Jun 2018 09:22:22	S	Std
7.5 - 1	57.5	29129	-	7.5000	7470A	PPB	04 Jun 2018 09:24:29	S	Std
10.0 - 1	510	38611	-	10.0000	7470A	PPB	04 Jun 2018 09:26:37	S	Std
ICV79 - 1	ICV79	15291	3.9123	-	7470A	PPB	04 Jun 2018 09:29:22	U	SMPL
ICB79 - 1	ICB79	378	0.0390	-	7470A	PPB	04 Jun 2018 09:31:28	U	SMPL
CCV30 - 1	CCV30	21129	5.4285	-	7470A	PPB	04 Jun 2018 09:33:35	U	SMPL
CCB30 - 1	CCB30	340	0.0291	-	7470A	PPB	04 Jun 2018 09:35:41	U	SMPL
CRA - 1	CRA	874	0.1678	-	7470A	PPB	04 Jun 2018 09:37:49	U	SMPL
HighStd - 1	HighStd	38390	9.9117	-	7470A	PPB	04 Jun 2018 09:39:55	U	SMPL
ChkStd - 1	ChkStd	27145	6.9911	-	7470A	PPB	04 Jun 2018 09:42:02	U	SMPL
PB109893BL - 1	PBW	241	0.0034	-	7470A	PPB	04 Jun 2018 09:46:37	U	SMPL
PB109893BS - 1	LCSW	14255	3.6432	-	7470A	PPB	04 Jun 2018 09:48:43	U	SMPL
PB109893TB - 1	PB109844TB	-29	-0.0667	-	7470A	PPB	04 Jun 2018 09:50:51	U	SMPL
J2670-08 - 1	TR-05-053118	178	-0.0130	-	7470A	PPB	04 Jun 2018 09:52:58	U	SMPL
J2672-10 - 1	BU-01-053118	233	0.0013	-	7470A	PPB	04 Jun 2018 09:55:06	U	SMPL
J2672-10DUP - 1	BU-01-053118DUP	216	-0.0031	-	7470A	PPB	04 Jun 2018 09:57:15	U	SMPL
J2672-10MS - 1	BU-01-053118MS	14560	3.7224	-	7470A	PPB	04 Jun 2018 09:59:21	U	SMPL
CCV31 - 1	CCV31	20037	5.1449	-	7470A	PPB	04 Jun 2018 10:01:28	U	SMPL
CCB31 - 1	CCB31	358	0.0338	-	7470A	PPB	04 Jun 2018 10:03:34	U	SMPL
J2672-10MSD - 1	BU-01-053118MSD	15014	3.8403	-	7470A	PPB	04 Jun 2018 10:05:42	U	SMPL
J2676-32 - 1	CL-02-053118-A	32	-0.0509	-	7470A	PPB	04 Jun 2018 10:07:48	U	SMPL
J2676-34 - 1	CL-02-053118-B	138	-0.0233	-	7470A	PPB	04 Jun 2018 10:09:55	U	SMPL
J2676-36 - 1	CL-02-053118-C	143	-0.0221	-	7470A	PPB	04 Jun 2018 10:12:02	U	SMPL
J2676-38 - 1	CL-02-053118-D	156	-0.0187	-	7470A	PPB	04 Jun 2018 10:14:10	U	SMPL
J2677-08 - 1	OK-03-053118	138	-0.0233	-	7470A	PPB	04 Jun 2018 10:16:17	U	SMPL
J3179-01 - 1	TYPE-A-TAGS	156	-0.0187	-	7470A	PPB	04 Jun 2018 10:18:25	U	SMPL
J3179-02 - 1	TYPE-B-TAGS	189	-0.0101	-	7470A	PPB	04 Jun 2018 10:20:33	U	SMPL
J3226-03 - 1	RS-RO-97-20	170	-0.0150	-	7470A	PPB	04 Jun 2018 10:22:42	U	SMPL
J3226-06 - 1	RS-RB47199RT	128	-0.0259	-	7470A	PPB	04 Jun 2018 10:24:49	U	SMPL
CCV32 - 1	CCV32	20095	5.1600	-	7470A	PPB	04 Jun 2018 10:26:56	U	SMPL
CCB32 - 1	CCB32	365	0.0356	-	7470A	PPB	04 Jun 2018 10:29:03	U	SMPL
J3227-03 - 1	RO-37426-30-CHERI-LANE	123	-0.0272	-	7470A	PPB	04 Jun 2018 10:31:11	U	SMPL
PB109894BL - 1	PBW	43	-0.0480	-	7470A	PPB	04 Jun 2018 10:33:18	U	SMPL
PB109894BS - 1	LCSW	15693	4.0167	-	7470A	PPB	04 Jun 2018 10:35:25	U	SMPL
PB109894TB - 1	PB109847TB	-74	-0.0784	-	7470A	PPB	04 Jun 2018 10:37:32	U	SMPL
J3237-04 - 1	AG-BASELINEWATER	32	-0.0509	-	7470A	PPB	04 Jun 2018 10:39:40	U	SMPL
J3237-04DUP - 1	AG-BASELINEWATERDUP	64	-0.0426	-	7470A	PPB	04 Jun 2018 10:41:47	U	SMPL
J3237-04MS - 1	AG-BASELINEWATERMS	14286	3.6512	-	7470A	PPB	04 Jun 2018 10:43:56	U	SMPL
J3237-04MSD - 1	AG-BASELINEWATERMSD	14613	3.7362	-	7470A	PPB	04 Jun 2018 10:46:04	U	SMPL
PB109895BL - 1	PBW	25	-0.0527	-	7470A	PPB	04 Jun 2018 10:48:12	U	SMPL
PB109895BS - 1	LCSW	14775	3.7782	-	7470A	PPB	04 Jun 2018 10:50:19	U	SMPL
CCV33 - 1	CCV33	20262	5.2034	-	7470A	PPB	04 Jun 2018 10:52:26	U	SMPL
CCB33 - 1	CCB33	-24	-0.0654	-	7470A	PPB	04 Jun 2018 10:54:33	U	SMPL
J3237-03 - 1	AG-BASELINEWATER	6185	1.5472	-	7470A	PPB	04 Jun 2018 10:56:40	U	SMPL
J3237-03DUP - 1	AG-BASELINEWATERDUP	5161	1.2813	-	7470A	PPB	04 Jun 2018 10:58:48	U	SMPL
J3237-03MS - 1	AG-BASELINEWATERMS	19611	5.0343	-	7470A	PPB	04 Jun 2018 11:03:52	U	SMPL
J3237-03MSD - 1	AG-BASELINEWATERMSD	20547	5.2774	-	7470A	PPB	04 Jun 2018 11:05:59	U	SMPL
J3273-01 - 1	GROUND-WATER	-141	-0.0958	-	7470A	PPB	04 Jun 2018 11:08:06	U	SMPL
CCV34 - 1	CCV34	20473	5.2582	-	7470A	PPB	04 Jun 2018 11:10:14	U	SMPL
CCB34 - 1	CCB34	-29	-0.0667	-	7470A	PPB	04 Jun 2018 11:12:22	U	SMPL
J2672-10LX5 - 1		159	-0.0179	-	7470A	PPB	04 Jun 2018 11:14:31	U	SMPL
J2672-10A - 1		15566	3.9837	-	7470A	PPB	04 Jun 2018 11:16:39	U	SMPL
J3237-04LX5 - 1		8	-0.0571	-	7470A	PPB	04 Jun 2018 11:18:47	U	SMPL
J3237-04A - 1		14385	3.6770	-	7470A	PPB	04 Jun 2018 11:20:53	U	SMPL
J3237-03LX5 - 1		1177	0.2465	-	7470A	PPB	04 Jun 2018 11:23:01	U	SMPL
J3237-03A - 1		21030	5.4028	-	7470A	PPB	04 Jun 2018 11:25:08	U	SMPL
CCV35 - 1	CCV35	19737	5.0670	-	7470A	PPB	04 Jun 2018 11:27:15	U	SMPL
CCB35 - 1	CCB35	287	0.0153	-	7470A	PPB	04 Jun 2018 11:29:24	U	SMPL