



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95785

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/01/2018	17:15
2	CAL2	1	Water			21.4	7.01	06/01/2018	17:17
3	CAL3	1	Water			21.6	10.03	06/01/2018	17:19
4	ICV	1	Water			21.3	7.00	06/01/2018	17:21
5	CCV1	1	Water			21.5	2.00	06/01/2018	17:23
6	J3273-01	1	Water			21.1	7.00	06/01/2018	17:25
7	J3273-01DUP	1	Water			21.2	7.01	06/01/2018	17:27
8	CCV2	1	Water			21.6	12.00	06/01/2018	17:29

LB 95785

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH060118

WorkList ID : 112883

Date : 6/1/2018 4:50:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/01/2018	Water	J3273-01	pH	Cool 4 deg C	PSEG01	I53	GROUND-WATER	06/01/2018	9040C

Date/Time 06/01/18 5:00 PM
Received by: SP
Relinquished by: CP

Date/Time 06/01/18 5:30 PM
Received by: CP
Relinquished by: SP