



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB95788
Analysis Date: 06/04/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 06/04/2018
Extraction IN Time: 09:45
Extraction OUT Time: 10:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB95788BL	LB95788BL	WATER	2	1000	100	3.0169	3.0169	3.02	3.0169	3.0169	0.0000	0
2	LB95788BS	LB95788BS	WATER	2	1000	100	3.1022	3.1022	3.01	3.1209	3.1209	0.0187	18.7
3	LB95788BSD	LB95788BSD	WATER	2	1000	100	3.1474	3.1474	3.03	3.1663	3.1663	0.0189	18.9
4	J2923-05	LOQ-NPW-Precision-Bias	WATER	2	1000	100	3.1463	3.1463	3.04	3.1501	3.1501	0.0038	3.8
5	J2923-06	LOQ-NPW-Precision-Bias	WATER	2	1000	100	3.0959	3.0959	3.06	3.0998	3.0998	0.0039	3.9
6	J2923-07	LOQ-NPW-Precision-Bias	WATER	2	1000	100	3.0751	3.0751	3.05	3.0792	3.0792	0.0041	4.1
7	J2923-08	LOQ-NPW-Precision-Bias	WATER	2	1000	100	3.0856	3.0856	3.02	3.0899	3.0899	0.0043	4.3
8	J3237-03	AG-BASELINEWATER	WATER	2	1000	100	3.1179	3.1179	30.00	3.3204	3.3204	0.2025	202.5
9	J3273-01	GROUND-WATER	WATER	2	1000	100	3.1235	3.1235	3.05	3.1241	3.1241	0.0006	0.6



QC Batch# LB95788 Test: TPH Analysis Date: 06/04/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2374
pH Paper 0-14	W2373
Sodium Sulfate	EP1853
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>12:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>10:15</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>12:45</u>
		Out Time:	<u>12:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>13:46</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>13:20</u>		
Bal Check Time:	<u>14:40</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>14:15</u>
		Out Time:	<u>13:45</u>		

LB45788

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J3277

WorkList ID : 112932

Date : 6/4/2018 9:32:05 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/22/2018	Water	J2923-05	TPH	Conc H2SO4 tc	CHEM02	QA Office	LOQ-NPW-Precision-Bias-1	05/15/2018	1664A
05/22/2018	Water	J2923-06	TPH	Conc H2SO4 tc	CHEM02	QA Office	LOQ-NPW-Precision-Bias-2	05/15/2018	1664A
05/22/2018	Water	J2923-07	TPH	Conc H2SO4 tc	CHEM02	QA Office	LOQ-NPW-Precision-Bias-3	05/15/2018	1664A
05/22/2018	Water	J2923-08	TPH	Conc H2SO4 tc	CHEM02	QA Office	LOQ-NPW-Precision-Bias-4	05/15/2018	1664A
06/05/2018	Water	J3237-03 -12	TPH	Conc H2SO4 tc	PSEG01	I43	AG-BASELINEWATER	05/31/2018	1664A
06/06/2018	Water	J3273-01 -17	TPH	Conc H2SO4 tc	PSEG01	I53	GROUND-WATER	06/01/2018	1664A

Date/Time 06/04/18 9:30 AM
 Received by: JD
 Relinquished by: CD

Date/Time 06/04/18 11:35 AM
 Received by: SC
 Relinquished by: JD