



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95790

Ambient Barometric Pressure (mmHg): 750

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	81.2	06/04/2018	09:10
2	J3257-02		1	62.0	143.6	144.2	06/04/2018	09:40
3	J3257-02DUP		1	62.0	143.6	144.2	06/04/2018	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06