



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95799

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/04/2018	14:00
2	CAL2	1	Water			21.4	7.00	06/04/2018	14:02
3	CAL3	1	Water			21.5	10.03	06/04/2018	14:04
4	ICV	1	Water			21.5	7.00	06/04/2018	14:06
5	CCV1	1	Water			21.4	2.00	06/04/2018	14:08
6	J3306-03	1	Water			14.8	3.96	06/04/2018	14:18
7	J3306-03DUP	1	Water			14.9	3.97	06/04/2018	14:20
8	CCV2	1	Water			21.5	12.00	06/04/2018	14:22

WORKLIST(Hardcopy Internal Chain)

LB 95799

WorkList Name : pH060418-1

WorkList ID : 112950

Date : 6/4/2018 1:42:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/01/2018	Water	J3306-03	pH	Cool 4 deg C	ENV176	P43	A&M-WATER	06/01/2018	9040C

Date/Time 06/04/18 1:45 PM
 Received by: sp
 Relinquished by: CP

Date/Time 06/04/18 2:35 PM
 Received by: CP
 Relinquished by: sp