



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95800

Ambient Barometric Pressure (mmHg): 750

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	83.9	06/04/2018	10:55
2	J3273-02		1	>100.0	>212.0	>212.0	06/04/2018	11:25
3	J3273-02DUP		1	>100.0	>212.0	>212.0	06/04/2018	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.33

LB95800

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP2-060418

WorkList ID : 112923

Date : 6/4/2018 8:19:37 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2018	Water	J3273-02	C Flash Point	Cool 4 deg C	PSEG01	I53	GROUND-WATER	06/01/2018	1010A

Date/Time 06/04/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 06/04/18 11:55 AM
Received by: CP
Relinquished by: KS