


245.1

LB95813
INSTRUMENT ID: CV2

Linear

 μ Abs.:

3511E



A= 0.0000e+000

B= 2.8440e-004

C= 6.0044e-003

Rho= 0.9999851

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-9	0.000	-9				
0.05	0.050	0.053	0.003	166	0.0 %	166				
0.2	0.200	0.189	-0.011	642	0.0 %	642				
2.5	2.500	2.523	0.023	8849	0.0 %	8849				
5.0	5.000	4.962	-0.038	17427	0.0 %	17427				
7.5	7.500	7.527	0.027	26446	0.0 %	26446				
10.0	10.000	9.993	-0.007	35116	0.0 %	35116				

LB95813 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 04 Jun 2018 17:29:40

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.0000245.1	PPB	04 Jun 2018 17:31:15	S	Std
0.05 - 1	50.05	166	-	0.0500245.1	PPB	04 Jun 2018 17:33:23	S	Std
0.2 - 1	50.2	642	-	0.2000245.1	PPB	04 Jun 2018 17:35:32	S	Std
2.5 - 1	52.5	8849	-	2.5000245.1	PPB	04 Jun 2018 17:37:40	S	Std
5.0 - 1	55	17427	-	5.0000245.1	PPB	04 Jun 2018 17:39:48	S	Std
7.5 - 1	57.5	26446	-	7.5000245.1	PPB	04 Jun 2018 17:41:55	S	Std
10.0 - 1	57.5	35116	-	10.0000245.1	PPB	04 Jun 2018 17:44:03	S	Std
ICV81 - 1	ICV81	14330	4.0814	-245.1	PPB	04 Jun 2018 17:49:13	U	SMPL
ICB81 - 1	ICB81	-123	-0.0290	-245.1	PPB	04 Jun 2018 17:51:19	U	SMPL
CCV42 - 1	CCV42	17447	4.9679	-245.1	PPB	04 Jun 2018 17:53:25	U	SMPL
CCB42 - 1	CCB42	-185	-0.0466	-245.1	PPB	04 Jun 2018 17:55:32	U	SMPL
CRA - 1	CRA	128	0.0424	-245.1	PPB	04 Jun 2018 17:57:40	U	SMPL
HighStd - 1	HighStd	35696	10.1578	-245.1	PPB	04 Jun 2018 17:59:46	U	SMPL
ChkStd - 1	ChkStd	23465	6.6793	-245.1	PPB	04 Jun 2018 18:01:53	U	SMPL
PB109896BL - 1	PBW	-156	-0.0384	-245.1	PPB	04 Jun 2018 18:04:02	U	SMPL
PB109896BS - 1	LCSW	14277	4.0663	-245.1	PPB	04 Jun 2018 18:06:11	U	SMPL
J3238-01 - 1	053018-DBRI-F-D-B	13	0.0097	-245.1	PPB	04 Jun 2018 18:08:19	U	SMPL
J3238-02 - 1	053018-DBRI-F-D-M	108	0.0367	-245.1	PPB	04 Jun 2018 18:10:27	U	SMPL
J3238-03 - 1	053018-DBRI-F-D-S	112	0.0379	-245.1	PPB	04 Jun 2018 18:12:35	U	SMPL
J3238-04 - 1	053018-DBRI-F-U-B	191	0.0603	-245.1	PPB	04 Jun 2018 18:14:44	U	SMPL
J3238-05 - 1	053018-DBRI-F-U-M	290	0.0885	-245.1	PPB	04 Jun 2018 18:16:50	U	SMPL
CCV43 - 1	CCV43	17316	4.9306	-245.1	PPB	04 Jun 2018 18:18:57	U	SMPL
CCB43 - 1	CCB43	-181	-0.0455	-245.1	PPB	04 Jun 2018 18:21:04	U	SMPL
J3238-06 - 1	053018-DBRI-F-U-S	223	0.0694	-245.1	PPB	04 Jun 2018 18:23:12	U	SMPL
J3238-07 - 1	053018-JETT-F-D-B	284	0.0868	-245.1	PPB	04 Jun 2018 18:25:18	U	SMPL
J3238-08 - 1	053018-JETT-F-D-M	492	0.1459	-245.1	PPB	04 Jun 2018 18:27:25	U	SMPL
J3238-09 - 1	053018-JETT-F-D-S	1166	0.3376	-245.1	PPB	04 Jun 2018 18:29:32	U	SMPL
J3238-10 - 1	053018-JETT-F-U-B	83	0.0296	-245.1	PPB	04 Jun 2018 18:31:39	U	SMPL
J3238-11 - 1	053018-JETT-F-U-M	106	0.0362	-245.1	PPB	04 Jun 2018 18:33:47	U	SMPL
J3238-12 - 1	053018-JETT-F-U-S	79	0.0285	-245.1	PPB	04 Jun 2018 18:35:54	U	SMPL
J3238-13 - 1	053018-DBRI-E-D-M	87	0.0307	-245.1	PPB	04 Jun 2018 18:38:03	U	SMPL
J3238-14 - 1	053018-DBRI-E-U-M	111	0.0376	-245.1	PPB	04 Jun 2018 18:40:11	U	SMPL
J3238-14DUP - 1	053018-DBRI-E-U-MDUP	106	0.0362	-245.1	PPB	04 Jun 2018 18:42:18	U	SMPL
CCV44 - 1	CCV44	16813	4.7875	-245.1	PPB	04 Jun 2018 18:44:25	U	SMPL
CCB44 - 1	CCB44	-173	-0.0432	-245.1	PPB	04 Jun 2018 18:46:32	U	SMPL
J3238-14MS - 1	053018-DBRI-E-U-MMS	17534	4.9926	-245.1	PPB	04 Jun 2018 18:48:41	U	SMPL
J3238-14MSD - 1	053018-DBRI-E-U-MMSD	17460	4.9716	-245.1	PPB	04 Jun 2018 18:50:48	U	SMPL
J3238-14LX5 - 1		-152	-0.0372	-245.1	PPB	04 Jun 2018 18:52:55	U	SMPL
J3238-14A - 1		15114	4.3044	-245.1	PPB	04 Jun 2018 18:55:02	U	SMPL
CCV45 - 1	CCV45	17151	4.8837	-245.1	PPB	04 Jun 2018 18:57:10	U	SMPL
CCB45 - 1	CCB45	-153	-0.0375	-245.1	PPB	04 Jun 2018 18:59:17	U	SMPL