


7470A

LB95821
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38441



A= 0.0000e+000

B= 2.5979e-004

C= -2.5127e-002

Rho= 0.9999672

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	38	0.000	38				
0.2	0.200	0.172	-0.028	759	0.0 %	759				
2.5	2.500	2.542	0.042	9880	0.0 %	9880				
5.0	5.000	5.008	0.008	19373	0.0 %	19373				
7.5	7.500	7.533	0.033	29092	0.0 %	29092				
10.0	10.000	9.961	-0.039	38441	0.0 %	38441				

LB95821 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jun 2018 11:29:28

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	7470A	PPB	05 Jun 2018 11:32:15	S	Std
0.2 - 1	50.2	759	-	0.2000	7470A	PPB	05 Jun 2018 11:34:23	S	Std
2.5 - 1	52.5	9880	-	2.5000	7470A	PPB	05 Jun 2018 11:36:31	S	Std
5.0 - 1	55	19373	-	5.0000	7470A	PPB	05 Jun 2018 11:38:39	S	Std
7.5 - 1	57.5	29092	-	7.5000	7470A	PPB	05 Jun 2018 11:43:14	S	Std
10.0 - 1	58	38441	-	10.0000	7470A	PPB	05 Jun 2018 11:45:20	S	Std
ICV82 - 1	ICV82	14266	3.6810	-7470A	PPB	PPB	05 Jun 2018 11:48:14	U	SMPL
ICB82 - 1	ICB82	-146	-0.0631	-7470A	PPB	PPB	05 Jun 2018 11:50:20	U	SMPL
CCV46 - 1	CCV46	18782	4.8542	-7470A	PPB	PPB	05 Jun 2018 11:52:27	U	SMPL
CCB46 - 1	CCB46	-138	-0.0610	-7470A	PPB	PPB	05 Jun 2018 11:54:33	U	SMPL
CRA - 1	CRA	743	0.1679	-7470A	PPB	PPB	05 Jun 2018 11:56:40	U	SMPL
HighStd - 1	HighStd	38746	10.0406	-7470A	PPB	PPB	05 Jun 2018 11:58:47	U	SMPL
ChkStd - 1	ChkStd	26269	6.7992	-7470A	PPB	PPB	05 Jun 2018 12:00:54	U	SMPL
PB109950BL - 1	PBW	-199	-0.0768	-7470A	PPB	PPB	05 Jun 2018 12:03:02	U	SMPL
PB109950BS - 1	LCSW	13256	3.4186	-7470A	PPB	PPB	05 Jun 2018 12:05:12	U	SMPL
PB109950TB - 1	PB109899TB	-96	-0.0501	-7470A	PPB	PPB	05 Jun 2018 12:07:19	U	SMPL
J3239-02 - 1	AU-05-060118-A	31	-0.0171	-7470A	PPB	PPB	05 Jun 2018 12:09:27	U	SMPL
J3239-02DUP - 1	AU-05-060118-ADUP	18	-0.0205	-7470A	PPB	PPB	05 Jun 2018 12:11:35	U	SMPL
J3239-02MS - 1	AU-05-060118-AMS	14692	3.7917	-7470A	PPB	PPB	05 Jun 2018 12:13:44	U	SMPL
J3239-02MSD - 1	AU-05-060118-AMSD	13880	3.5807	-7470A	PPB	PPB	05 Jun 2018 12:15:50	U	SMPL
CCV47 - 1	CCV47	18513	4.7843	-7470A	PPB	PPB	05 Jun 2018 12:17:57	U	SMPL
CCB47 - 1	CCB47	-153	-0.0649	-7470A	PPB	PPB	05 Jun 2018 12:20:05	U	SMPL
J3239-04 - 1	AU-05-060118-B	-7	-0.0269	-7470A	PPB	PPB	05 Jun 2018 12:22:12	U	SMPL
J3239-06 - 1	AU-05-060118-C	21	-0.0197	-7470A	PPB	PPB	05 Jun 2018 12:24:19	U	SMPL
J3239-08 - 1	AU-05-060118-D	22	-0.0194	-7470A	PPB	PPB	05 Jun 2018 12:26:26	U	SMPL
J3259-03 - 1	RS-RO-274318	23	-0.0192	-7470A	PPB	PPB	05 Jun 2018 12:28:33	U	SMPL
J3262-03 - 1	RS-RO-274319	22	-0.0194	-7470A	PPB	PPB	05 Jun 2018 12:30:40	U	SMPL
J3263-03 - 1	RS-RO-124022	35	-0.0160	-7470A	PPB	PPB	05 Jun 2018 12:32:48	U	SMPL
J3266-03 - 1	RS-RB-36793RT	44	-0.0137	-7470A	PPB	PPB	05 Jun 2018 12:34:56	U	SMPL
J3271-02 - 1	TP-7A-WC-1	27	-0.0181	-7470A	PPB	PPB	05 Jun 2018 12:37:04	U	SMPL
J3271-04 - 1	TP-7A-WC-2	10	-0.0225	-7470A	PPB	PPB	05 Jun 2018 12:39:12	U	SMPL
J3271-06 - 1	TP-7A-WC-3	17	-0.0207	-7470A	PPB	PPB	05 Jun 2018 12:41:19	U	SMPL
CCV48 - 1	CCV48	18885	4.8809	-7470A	PPB	PPB	05 Jun 2018 12:43:26	U	SMPL
CCB48 - 1	CCB48	-130	-0.0589	-7470A	PPB	PPB	05 Jun 2018 12:45:33	U	SMPL
J3271-08 - 1	TP-7A-WC-4	8	-0.0230	-7470A	PPB	PPB	05 Jun 2018 12:47:41	U	SMPL
J3272-02 - 1	RBR-45901-RBR-47184-RBR-47183	22	-0.0194	-7470A	PPB	PPB	05 Jun 2018 12:49:48	U	SMPL
J3276-05 - 1	ES-STONE	36	-0.0158	-7470A	PPB	PPB	05 Jun 2018 12:51:55	U	SMPL
J3283-01 - 1	SS-OILY-DEBRIS	15	-0.0212	-7470A	PPB	PPB	05 Jun 2018 12:54:02	U	SMPL
J3286-02 - 1	TP-6A-WC-1	2	-0.0246	-7470A	PPB	PPB	05 Jun 2018 12:56:09	U	SMPL
J3286-04 - 1	TP-6A-WC-2	24	-0.0189	-7470A	PPB	PPB	05 Jun 2018 12:58:17	U	SMPL
J3286-06 - 1	TP-6A-WC-3	41	-0.0145	-7470A	PPB	PPB	05 Jun 2018 13:00:25	U	SMPL
J3286-08 - 1	TP-6A-WC-4	-4	-0.0262	-7470A	PPB	PPB	05 Jun 2018 13:02:33	U	SMPL
J3286-10 - 1	TP-6A-WC-5	109	0.0032	-7470A	PPB	PPB	05 Jun 2018 13:04:42	U	SMPL
PB109951BL - 1	PBW	26	-0.0184	-7470A	PPB	PPB	05 Jun 2018 13:06:48	U	SMPL
CCV49 - 1	CCV49	18347	4.7412	-7470A	PPB	PPB	05 Jun 2018 13:08:55	U	SMPL
CCB49 - 1	CCB49	-129	-0.0586	-7470A	PPB	PPB	05 Jun 2018 13:11:03	U	SMPL
PB109951BS - 1	LCSW	14427	3.7228	-7470A	PPB	PPB	05 Jun 2018 13:13:11	U	SMPL
PB109951TB - 1	PB109901TB	-54	-0.0392	-7470A	PPB	PPB	05 Jun 2018 13:15:18	U	SMPL
J3243-02 - 1	NB-08-060118-A	24	-0.0189	-7470A	PPB	PPB	05 Jun 2018 13:17:25	U	SMPL
J3243-02DUP - 1	NB-08-060118-ADUP	34	-0.0163	-7470A	PPB	PPB	05 Jun 2018 13:19:32	U	SMPL
J3243-02MS - 1	NB-08-060118-AMS	13749	3.5467	-7470A	PPB	PPB	05 Jun 2018 13:21:39	U	SMPL
J3243-02MSD - 1	NB-08-060118-AMSD	14267	3.6812	-7470A	PPB	PPB	05 Jun 2018 13:23:47	U	SMPL
J3243-04 - 1	NB-08-060118-B	-73	-0.0441	-7470A	PPB	PPB	05 Jun 2018 13:25:55	U	SMPL
J3277-04 - 1	OILY-STONES	9	-0.0228	-7470A	PPB	PPB	05 Jun 2018 13:28:03	U	SMPL
PB109952BL - 1	PBW	0	-0.0251	-7470A	PPB	PPB	05 Jun 2018 13:30:11	U	SMPL
PB109952BS - 1	LCSW	13635	3.5171	-7470A	PPB	PPB	05 Jun 2018 13:32:18	U	SMPL
CCV50 - 1	CCV50	19375	5.0082	-7470A	PPB	PPB	05 Jun 2018 13:34:26	U	SMPL
CCB50 - 1	CCB50	-141	-0.0618	-7470A	PPB	PPB	05 Jun 2018 13:36:33	U	SMPL
PB109952TB - 1	PB109903TB	18	-0.0205	-7470A	PPB	PPB	05 Jun 2018 13:38:41	U	SMPL
J3273-02 - 1	GROUND-WATER	4	-0.0241	-7470A	PPB	PPB	05 Jun 2018 13:40:48	U	SMPL
J3273-02DUP - 1	GROUND-WATERDUP	10	-0.0225	-7470A	PPB	PPB	05 Jun 2018 13:42:56	U	SMPL
J3273-02MS - 1	GROUND-WATERMS	13642	3.5189	-7470A	PPB	PPB	05 Jun 2018 13:45:02	U	SMPL
J3273-02MSD - 1	GROUND-WATERMSD	13964	3.6025	-7470A	PPB	PPB	05 Jun 2018 13:47:10	U	SMPL
CCV51 - 1	CCV51	19399	5.0145	-7470A	PPB	PPB	05 Jun 2018 13:49:18	U	SMPL
CCB51 - 1	CCB51	-121	-0.0566	-7470A	PPB	PPB	05 Jun 2018 13:51:25	U	SMPL
J3239-02LX5 - 1		22	-0.0194	-7470A	PPB	PPB	05 Jun 2018 13:53:34	U	SMPL
J3239-02A - 1		14091	3.6355	-7470A	PPB	PPB	05 Jun 2018 13:55:43	U	SMPL
J3243-02LX5 - 1		-77	-0.0451	-7470A	PPB	PPB	05 Jun 2018 13:57:50	U	SMPL
J3243-02A - 1		14687	3.7904	-7470A	PPB	PPB	05 Jun 2018 13:59:58	U	SMPL
J3273-02LX5 - 1		-104	-0.0521	-7470A	PPB	PPB	05 Jun 2018 14:02:06	U	SMPL
J3273-02A - 1		14166	3.6550	-7470A	PPB	PPB	05 Jun 2018 14:04:14	U	SMPL
CCV52 - 1	CCV52	18479	4.7755	-7470A	PPB	PPB	05 Jun 2018 14:06:21	U	SMPL
CCB52 - 1	CCB52	-143	-0.0623	-7470A	PPB	PPB	05 Jun 2018 14:08:28	U	SMPL