

245.1


LB95825

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34501



A= 0.0000e+000

B= 2.9227e-004

C= 2.6251e-003

Rho= 0.9998849

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	23	0.000	23				
0.05	0.050	0.056	0.006	184	0.0 %	184				
0.2	0.200	0.203	0.003	686	0.0 %	686				
2.5	2.500	2.547	0.047	8705	0.0 %	8705				
5.0	5.000	4.945	-0.055	16911	0.0 %	16911				
7.5	7.500	7.403	-0.097	25320	0.0 %	25320				
10.0	10.000	10.086	0.086	34501	0.0 %	34501				

LB95825 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 05 Jun 2018 15:09:08					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	23	-	0.0000	245.1	PPB	05 Jun 2018 15:11:03	S	Std
0.05 - 1	50.05	184	-	0.0500	245.1	PPB	05 Jun 2018 15:13:11	S	Std
0.2 - 1	50.2	686	-	0.2000	245.1	PPB	05 Jun 2018 15:15:19	S	Std
2.5 - 1	52.5	8705	-	2.5000	245.1	PPB	05 Jun 2018 15:17:27	S	Std
5.0 - 1	55	16911	-	5.0000	245.1	PPB	05 Jun 2018 15:19:35	S	Std
7.5 - 1	57.5	25320	-	7.5000	245.1	PPB	05 Jun 2018 15:21:42	S	Std
10.0 - 1	510	34501	-	10.0000	245.1	PPB	05 Jun 2018 15:23:50	S	Std
ICV83 - 1	ICV83	13039	3.8135	-245.1		PPB	05 Jun 2018 15:26:38	U	SMPL
ICB83 - 1	ICB83	-140	-0.0383	-245.1		PPB	05 Jun 2018 15:28:44	U	SMPL
CCV53 - 1	CCV53	17129	5.0089	-245.1		PPB	05 Jun 2018 15:30:51	U	SMPL
CCB53 - 1	CCB53	-157	-0.0433	-245.1		PPB	05 Jun 2018 15:32:58	U	SMPL
CRA - 1	CRA	151	0.0468	-245.1		PPB	05 Jun 2018 15:35:05	U	SMPL
HighStd - 1	HighStd	34433	10.0663	-245.1		PPB	05 Jun 2018 15:37:12	U	SMPL
ChkStd - 1	ChkStd	23711	6.9326	-245.1		PPB	05 Jun 2018 15:39:18	U	SMPL
PB109976BL - 1	PBW	-232	-0.0652	-245.1		PPB	05 Jun 2018 15:41:28	U	SMPL
PB109976BS - 1	LCSW	14497	4.2397	-245.1		PPB	05 Jun 2018 15:43:37	U	SMPL
J3238-04 - 1	053018-DBRI-F-U-B	-78	-0.0202	-245.1		PPB	05 Jun 2018 15:45:44	U	SMPL
J3238-05 - 1	053018-DBRI-F-U-M	-27	-0.0053	-245.1		PPB	05 Jun 2018 15:47:53	U	SMPL
J3238-06 - 1	053018-DBRI-F-U-S	-1	0.0023	-245.1		PPB	05 Jun 2018 15:50:01	U	SMPL
J3238-07 - 1	053018-JETT-F-D-B	27	0.0105	-245.1		PPB	05 Jun 2018 15:52:09	U	SMPL
J3238-08 - 1	053018-JETT-F-D-M	38	0.0137	-245.1		PPB	05 Jun 2018 15:54:16	U	SMPL
CCV54 - 1	CCV54	17043	4.9838	-245.1		PPB	05 Jun 2018 15:56:23	U	SMPL
CCB54 - 1	CCB54	-160	-0.0441	-245.1		PPB	05 Jun 2018 15:58:30	U	SMPL
J3238-09 - 1	053018-JETT-F-D-S	15	0.0070	-245.1		PPB	05 Jun 2018 16:00:38	U	SMPL
J3281-01 - 1	053118-DBRC-F-D-M	2	0.0032	-245.1		PPB	05 Jun 2018 16:02:45	U	SMPL
J3281-02 - 1	053118-DBRC-F-U-M	15	0.0070	-245.1		PPB	05 Jun 2018 16:04:52	U	SMPL
J3281-03 - 1	053118-JETT-F-D-B	14	0.0067	-245.1		PPB	05 Jun 2018 16:06:59	U	SMPL
J3281-04 - 1	053118-JETT-F-D-M	20	0.0085	-245.1		PPB	05 Jun 2018 16:09:06	U	SMPL
J3281-05 - 1	053118-JETT-F-D-S	17	0.0076	-245.1		PPB	05 Jun 2018 16:11:14	U	SMPL
J3281-06 - 1	053118-JETT-F-U-B	40	0.0143	-245.1		PPB	05 Jun 2018 16:13:22	U	SMPL
J3281-07 - 1	053118-JETT-F-U-M	29	0.0111	-245.1		PPB	05 Jun 2018 16:15:30	U	SMPL
J3281-08 - 1	053118-JETT-F-U-S	33	0.0123	-245.1		PPB	05 Jun 2018 16:17:39	U	SMPL
J3281-09 - 1	053118-DBRI-F-D-B	33	0.0123	-245.1		PPB	05 Jun 2018 16:19:46	U	SMPL
CCV55 - 1	CCV55	17139	5.0118	-245.1		PPB	05 Jun 2018 16:21:53	U	SMPL
CCB55 - 1	CCB55	-145	-0.0398	-245.1		PPB	05 Jun 2018 16:24:00	U	SMPL
J3281-10 - 1	053118-DBRI-F-D-S	-2	0.0020	-245.1		PPB	05 Jun 2018 16:26:09	U	SMPL
J3281-11 - 1	053118-DBRI-F-U-B	6	0.0044	-245.1		PPB	05 Jun 2018 16:28:16	U	SMPL
J3281-12 - 1	053118-DBRI-F-U-S	15	0.0070	-245.1		PPB	05 Jun 2018 16:30:23	U	SMPL
J3281-12DUP - 1	053118-DBRI-F-U-SDUP	19	0.0082	-245.1		PPB	05 Jun 2018 16:32:30	U	SMPL
J3281-12MS - 1	053118-DBRI-F-U-SMS	16127	4.7161	-245.1		PPB	05 Jun 2018 16:34:37	U	SMPL
J3281-12MSD - 1	053118-DBRI-F-U-SMSD	15704	4.5924	-245.1		PPB	05 Jun 2018 16:36:45	U	SMPL
J3281-12LX5 - 1		-22	-0.0038	-245.1		PPB	05 Jun 2018 16:44:36	U	SMPL
J3281-12A - 1		16054	4.6947	-245.1		PPB	05 Jun 2018 16:46:44	U	SMPL
CCV56 - 1	CCV56	16941	4.9540	-245.1		PPB	05 Jun 2018 16:48:52	U	SMPL
CCB56 - 1	CCB56	-151	-0.0415	-245.1		PPB	05 Jun 2018 16:50:59	U	SMPL