



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95832

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/05/2018	16:10
2	CAL2	1	Water			21.7	7.00	06/05/2018	16:12
3	CAL3	1	Water			21.4	10.04	06/05/2018	16:14
4	ICV	1	Water			21.4	7.00	06/05/2018	16:16
5	CCV1	1	Water			21.4	2.00	06/05/2018	16:20
6	J3241-02	1	Solid	20.02	20	23.1	9.29	06/05/2018	16:22
7	J3241-02DUP	1	Solid	20.03	20	23.2	9.30	06/05/2018	16:24
8	J3241-04	1	Solid	20.04	20	23.2	9.18	06/05/2018	16:26
9	J3241-06	1	Solid	20.02	20	23.2	9.53	06/05/2018	16:28
10	J3251-02	1	Solid	20.05	20	23.0	8.96	06/05/2018	16:30
11	J3325-02	1	Solid	20.02	20	22.8	10.70	06/05/2018	16:32
12	J3326-03	1	Solid	20.03	20	22.7	8.58	06/05/2018	16:34
13	CCV2	1	Water			21.5	12.00	06/05/2018	16:36

WORKLIST(Hardcopy Internal Chain)

LB 95832

WorkList Name : corrosivity060518

WorkList ID : 113002

Date : 6/5/2018 3:27:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2018	Solid	J3241-02	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-06-060518-A	06/05/2018	9045C
06/05/2018	Solid	J3241-04	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-06-060518-B	06/05/2018	9045C
06/05/2018	Solid	J3241-06	Corrosivity	Cool 4 deg C	PSEG05	I63	HR-06-060518-C	06/05/2018	9045C
06/05/2018	Solid	J3251-02	Corrosivity	Cool 4 deg C	PSEG05	I63	SU-02-060518	06/05/2018	9045C
06/05/2018	Solid	J3325-02	Corrosivity	Cool 4 deg C	PSEG01	I63	PS-OILY-STONE-RO-156	06/05/2018	9045C
06/05/2018	Solid	J3326-03	Corrosivity	Cool 4 deg C	DIST01	P51	PATH-SS7-WC1-060518	06/05/2018	9045C

Date/Time 06/05/18 4:00 PM

Received by: SP

Relinquished by: CP

Date/Time 06/05/18 4:45 PM

Received by: CP

Relinquished by: SP