

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/06/2018

TEMP1 IN: 103 06/06/2018 07:10 TEMP OUT: 104 06/06/2018 08:20

TEMP2 IN: 104 06/06/2018 09:00 TEMP OUT: 104 06/06/2018 10:10

TEMP3 IN: 105 06/06/2018 11:30 TEMP OUT: 105 06/06/2018 13:05

TEMP4 IN: 105 06/06/2018 13:45 TEMP OUT: 105 06/06/2018 15:25

Run Number: LB95834

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95834BL	LB95834BL	1.4083	1.4083	100	1.4083	1.4083	1.4083	0.0000	0
2	LB95834BS	LB95834BS	1.4106	1.4106	100	1.4610	1.4610	1.4610	0.0504	504
3	J3321-01	060118-JETT-F-D-B	1.4012	1.4012	200	1.4203	1.4203	1.4203	0.0191	95.5
4	J3321-01DUP	060118-JETT-F-D-BDUP	1.4061	1.4061	200	1.4253	1.4253	1.4253	0.0192	96
5	J3321-02	060118-JETT-F-D-M	1.4019	1.4019	200	1.4190	1.4190	1.4190	0.0171	85.5
6	J3321-03	060118-JETT-F-D-S	1.4045	1.4045	400	1.4161	1.4161	1.4161	0.0116	29
7	J3321-04	060118-JETT-F-U-B	1.4039	1.4039	200	1.4337	1.4337	1.4337	0.0298	149
8	J3321-05	060118-JETT-F-U-M	1.4030	1.4030	200	1.4168	1.4168	1.4168	0.0138	69
9	J3321-06	060118-JETT-F-U-S	1.3928	1.3928	300	1.4008	1.4008	1.4008	0.0080	26.7
10	J3323-01	060418-JETT-E-D-B	1.4007	1.4007	250	1.4133	1.4133	1.4133	0.0126	50.4
11	J3323-02	060418-JETT-E-D-M	1.4107	1.4107	250	1.4170	1.4170	1.4170	0.0063	25.2
12	J3323-03	060418-JETT-E-D-S	1.4076	1.4076	400	1.4150	1.4150	1.4150	0.0074	18.5
13	J3323-04	060418-JETT-E-U-B	1.4024	1.4024	400	1.4178	1.4178	1.4178	0.0154	38.5
14	J3323-05	060418-JETT-E-U-M	1.4006	1.4006	500	1.4157	1.4157	1.4157	0.0151	30.2
15	J3323-06	060418-JETT-E-U-S	1.3937	1.3937	500	1.4006	1.4006	1.4006	0.0069	13.8

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/06/2018

TEMP1 IN: 103 06/06/2018 07:10 TEMP OUT: 104 06/06/2018 08:20

TEMP2 IN: 104 06/06/2018 09:00 TEMP OUT: 104 06/06/2018 10:10

TEMP3 IN: 105 06/06/2018 11:30 TEMP OUT: 105 06/06/2018 13:05

TEMP4 IN: 105 06/06/2018 13:45 TEMP OUT: 105 06/06/2018 15:25

Run Number: LB95834

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J3323-07	060418-DBRI-E-D-B	1.4080	1.4080	500	1.4211	1.4211	1.4211	0.0131	26.2
17	J3323-08	060418-DBRI-E-D-S	1.3959	1.3959	400	1.4058	1.4058	1.4058	0.0099	24.8
18	J3323-09	060418-DBRI-E-U-B	1.4051	1.4051	400	1.4158	1.4158	1.4158	0.0107	26.8
19	J3323-10	060418-DBRI-E-U-S	1.4037	1.4037	250	1.4082	1.4082	1.4082	0.0045	18
20	J3323-10DUP	060418-DBRI-E-U-SDUP	1.4017	1.4017	250	1.4061	1.4061	1.4061	0.0044	17.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS060618

WorkList ID : 113007

Date : 6/6/2018 7:15:10 AM

2895834

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2018	Water	J3321-01 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-D-B	06/01/2018	SM2540 D
06/08/2018	Water	J3321-02 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-D-M	06/01/2018	SM2540 D
06/08/2018	Water	J3321-03 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-D-S	06/01/2018	SM2540 D
06/08/2018	Water	J3321-04 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-U-B	06/01/2018	SM2540 D
06/08/2018	Water	J3321-05 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-U-M	06/01/2018	SM2540 D
06/08/2018	Water	J3321-06 C	TSS	Cool 4 deg C	HDR103	P21	060118-JETT-F-U-S	06/01/2018	SM2540 D
06/11/2018	Water	J3323-01 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-D-B	06/04/2018	SM2540 D
06/11/2018	Water	J3323-02 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-D-M	06/04/2018	SM2540 D
06/11/2018	Water	J3323-03 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-D-S	06/04/2018	SM2540 D
06/11/2018	Water	J3323-04 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-U-B	06/04/2018	SM2540 D
06/11/2018	Water	J3323-05 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-U-M	06/04/2018	SM2540 D
06/11/2018	Water	J3323-06 C	TSS	Cool 4 deg C	HDR103	P21	060418-JETT-E-U-S	06/04/2018	SM2540 D
06/11/2018	Water	J3323-07 C	TSS	Cool 4 deg C	HDR103	P21	060418-DBRI-E-D-B	06/04/2018	SM2540 D
06/11/2018	Water	J3323-08 C	TSS	Cool 4 deg C	HDR103	P21	060418-DBRI-E-D-S	06/04/2018	SM2540 D
06/11/2018	Water	J3323-09 C	TSS	Cool 4 deg C	HDR103	P21	060418-DBRI-E-U-B	06/04/2018	SM2540 D
06/11/2018	Water	J3323-10 C	TSS	Cool 4 deg C	HDR103	P21	060418-DBRI-E-U-S	06/04/2018	SM2540 D

Date/Time 06/06/18, 08:00
 Received by: AS
 Relinquished by: DD

Date/Time 06/06/18, 12:00
 Received by: AS
 Relinquished by: AS