


7471B

LB95844
INSTRUMENT ID: CV2

Linear

 μ Abs.:

4494E



A= 0.0000e+000

B= 2.2218e-004

C= -1.7801e-002

Rho= 0.9999372

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	15	0.000	15				
0.2	0.200	0.161	-0.039	806	0.0 %	806				
2.5	2.500	2.577	0.077	11679	0.0 %	11679				
5.0	5.000	4.975	-0.025	22473	0.0 %	22473				
7.5	7.500	7.532	0.032	33981	0.0 %	33981				
10.0	10.000	9.969	-0.031	44948	0.0 %	44948				

LB95844 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jun 2018 14:00:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	7471B	PPB	06 Jun 2018 14:06:25	S	Std
0.2 - 1	50.2	806	-	0.2000	7471B	PPB	06 Jun 2018 14:08:42	S	Std
2.5 - 1	52.5	11679	-	2.5000	7471B	PPB	06 Jun 2018 14:11:00	S	Std
5.0 - 1	55	22473	-	5.0000	7471B	PPB	06 Jun 2018 14:13:17	S	Std
7.5 - 1	57.5	33981	-	7.5000	7471B	PPB	06 Jun 2018 14:15:34	S	Std
10.0 - 1	510	44948	-	10.0000	7471B	PPB	06 Jun 2018 14:17:52	S	Std
ICV86 - 1	ICV86	17633	3.8999	-	7471B	PPB	06 Jun 2018 14:21:10	U	SMPL
ICB86 - 1	ICB86	-126	-0.0458	-	7471B	PPB	06 Jun 2018 14:23:25	U	SMPL
CCV65 - 1	CCV65	22457	4.9717	-	7471B	PPB	06 Jun 2018 14:25:41	U	SMPL
CCB65 - 1	CCB65	-142	-0.0494	-	7471B	PPB	06 Jun 2018 14:27:57	U	SMPL
CRA - 1	CRA	801	0.1602	-	7471B	PPB	06 Jun 2018 14:30:13	U	SMPL
HighStd - 1	HighStd	45472	10.0852	-	7471B	PPB	06 Jun 2018 14:32:29	U	SMPL
ChkStd - 1	ChkStd	31970	7.0853	-	7471B	PPB	06 Jun 2018 14:34:44	U	SMPL
PB110003BL - 1	PBS	-219	-0.0665	-	7471B	PPB	06 Jun 2018 14:37:02	U	SMPL
PB110003BS - 1	LCSS	14845	3.2805	-	7471B	PPB	06 Jun 2018 14:39:21	U	SMPL
J3241-01 - 1	HR-06-060518-A	2350	0.5043	-	7471B	PPB	06 Jun 2018 14:41:38	U	SMPL
J3241-01DUP - 1	HR-06-060518-ADUP	2215	0.4743	-	7471B	PPB	06 Jun 2018 14:43:55	U	SMPL
J3241-01MS - 1	HR-06-060518-AMS	19519	4.3189	-	7471B	PPB	06 Jun 2018 14:46:13	U	SMPL
J3241-01MSD - 1	HR-06-060518-AMSD	19048	4.2143	-	7471B	PPB	06 Jun 2018 14:48:30	U	SMPL
J3241-03 - 1	HR-06-060518-B	6067	1.3302	-	7471B	PPB	06 Jun 2018 14:50:46	U	SMPL
CCV66 - 1	CCV66	22626	5.0093	-	7471B	PPB	06 Jun 2018 14:53:02	U	SMPL
CCB66 - 1	CCB66	-157	-0.0527	-	7471B	PPB	06 Jun 2018 14:55:18	U	SMPL
J3241-05 - 1	HR-06-060518-C	1893	0.4028	-	7471B	PPB	06 Jun 2018 14:57:34	U	SMPL
J3251-01 - 1	SU-02-060518	1708	0.3617	-	7471B	PPB	06 Jun 2018 14:59:50	U	SMPL
J3324-01 - 1	EPE-109-1(0-5)	2788	0.6016	-	7471B	PPB	06 Jun 2018 15:02:06	U	SMPL
J3325-01 - 1	BS-OILY-STONE-RO-156	123	0.0095	-	7471B	PPB	06 Jun 2018 15:04:23	U	SMPL
J3326-02 - 1	PATH-SS7-WC1-060518	24610	5.4501	-	7471B	PPB	06 Jun 2018 15:06:39	U	SMPL
J3241-01LX5 - 1		270	0.0422	-	7471B	PPB	06 Jun 2018 15:08:56	U	SMPL
J3241-01A - 1		20725	4.5869	-	7471B	PPB	06 Jun 2018 15:11:13	U	SMPL
CCV67 - 1	CCV67	22564	4.9955	-	7471B	PPB	06 Jun 2018 15:13:31	U	SMPL
CCB67 - 1	CCB67	-97	-0.0394	-	7471B	PPB	06 Jun 2018 15:15:49	U	SMPL