



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95850

Slope : 97.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/06/2018	12:36
2	CAL2	1	Water			21.6	7.00	06/06/2018	12:38
3	CAL3	1	Water			21.4	10.03	06/06/2018	12:40
4	ICV	1	Water			21.5	7.00	06/06/2018	12:42
5	CCV1	1	Water			21.6	2.00	06/06/2018	12:44
6	J3338-01	1	Water			21.6	4.02	06/06/2018	12:46
7	J3338-01DUP	1	Water			21.7	4.03	06/06/2018	12:48
8	J3341-01	1	Water			21.6	4.92	06/06/2018	12:50
9	CCV2	1	Water			21.4	12.00	06/06/2018	12:52

LB 95850

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH060618 WorkList ID : 113035 Date : 6/6/2018 12:23:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2018	Water	J3338-01	pH	Cool 4 deg C	STER14	P51	MINI-CELL	06/04/2018	SM 4500-PH B
06/05/2018	Water	J3341-01	pH	Cool 4 deg C	STER14	P22	JS-SYSTEM	06/05/2018	SM 4500-PH B

Date/Time 06/06/18 12:30 PM
Received by: SP
Relinquished by: CP

Date/Time 06/06/18 12:55 PM
Received by: CP
Relinquished by: SP