



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95851

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/06/2018	16:25
2	CAL2	1	Water			21.6	7.00	06/06/2018	16:27
3	CAL3	1	Water			21.4	10.03	06/06/2018	16:29
4	ICV	1	Water			21.6	7.00	06/06/2018	16:31
5	CCV1	1	Water			21.7	2.00	06/06/2018	16:33
6	J3252-02	1	Solid	20.02	20	22.4	8.32	06/06/2018	16:35
7	J3252-02DUP	1	Solid	20.04	20	22.5	8.33	06/06/2018	16:37
8	J3344-04	1	Solid	20.03	20	22.9	10.44	06/06/2018	16:39
9	J3344-08	1	Solid	20.02	20	22.7	10.48	06/06/2018	16:41
10	J3344-12	1	Solid	20.02	20	23.0	9.74	06/06/2018	16:43
11	J3352-04	1	Solid	20.04	20	22.6	8.36	06/06/2018	16:45
12	CCV2	1	Water			21.5	12.00	06/06/2018	16:47

WORKLIST(Hardcopy Internal Chain)

LB 95851

WorkList Name : corrosivity060618

WorkList ID : 113040

Date : 6/6/2018 4:10:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2018	Solid	J3252-02	Corrosivity	Cool 4 deg C	PSEG05	I63	TR-04-060618	06/06/2018	9045C
06/06/2018	Solid	J3344-04	Corrosivity	Cool 4 deg C	PSEG01	I63	HS-YARD-1	06/06/2018	9045C
06/06/2018	Solid	J3344-08	Corrosivity	Cool 4 deg C	PSEG01	I63	HS-YARD-2	06/06/2018	9045C
06/06/2018	Solid	J3344-12	Corrosivity	Cool 4 deg C	PSEG01	I63	HS-HOUSE-WALL	06/06/2018	9045C
06/06/2018	Solid	J3352-04	Corrosivity	Cool 4 deg C	PSEG01	I63	PS-SOIL	06/06/2018	9045C

Date/Time 06/06/18 4:15 PM

Received by: SP

Relinquished by: CP

Date/Time 06/06/18 5:00 PM

Received by: CP

Relinquished by: SP