

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/7/2018 15:00

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.803	0.0	0.105	
ICB1	-0.594	0.0	0.004	
CCV1	246.593	0.0	0.259	
CCB1	-0.786	0.0	0.004	
PB110036BL	-0.787	0.0	0.004	
PB110036BS	98.778	0.0	0.106	
LOWPB110036	9.207	0.0	0.014	
HIGHPB110036	483.227	0.0	0.503	
J3252-01	-0.121	0.0	0.004	
J3344-01	4.738	0.0	0.009	
J3344-05	0.937	0.0	0.005	
J3344-09	2.492	0.0	0.007	
J3350-01	-0.899	0.0	0.003	
CCV2	246.030	0.0	0.258	
CCB2	-0.145	0.0	0.004	
J3354-02	-0.266	0.0	0.004	
J3354-02DUP	-0.326	0.0	0.004	
J3354-03	30.890	0.0	0.036	
J3354-04	35.097	0.0	0.041	
J3354-05	-0.148	0.0	0.004	
J3352-01	-0.213	0.0	0.004	
CCV3	259.927	0.0	0.273	
CCB3	-0.548	0.0	0.004	

N 23
Mean 65.691
SD 125.1879
CV% 190.57

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 07 14:30:38 2018

Thu Jun 07 14:59:45 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.2083	µg/l	6/7/2018 10:07:46	
5.0PPBCN	A	Total CN	P	4.229	µg/l	6/7/2018 10:07:47	
10PPBCN	A	Total CN	P	9.6063	µg/l	6/7/2018 10:07:48	
50PPBCN	A	Total CN	P	50.2503	µg/l	6/7/2018 10:07:49	
100PPBCN	A	Total CN	P	100.9506	µg/l	6/7/2018 10:07:50	
250PPBCN	A	Total CN	P	252.7433	µg/l	6/7/2018 10:07:51	
500PPBCN	A	Total CN	P	498.4288	µg/l	6/7/2018 10:07:52	
ICV1	S	Total CN	P	97.803	µg/l	6/7/2018 14:30:38	
ICB1	S	Total CN	P	-0.5944	µg/l	6/7/2018 14:30:39	
CCV1	S	Total CN	P	246.5933	µg/l	6/7/2018 14:30:40	
CCB1	S	Total CN	P	-0.7858	µg/l	6/7/2018 14:30:41	
PB110036BL	S	Total CN	P	-0.7873	µg/l	6/7/2018 14:30:42	
PB110036BS	S	Total CN	P	98.7783	µg/l	6/7/2018 14:30:43	
LOWPB110036	S	Total CN	P	9.2072	µg/l	6/7/2018 14:30:44	
HIGHPB110036	S	Total CN	P	483.2274	µg/l	6/7/2018 14:30:45	
J3252-01	S	Total CN	P	-0.1208	µg/l	6/7/2018 14:30:46	
J3344-01	S	Total CN	P	4.7383	µg/l	6/7/2018 14:30:47	
J3344-05	S	Total CN	P	0.9371	µg/l	6/7/2018 14:30:48	
J3344-09	S	Total CN	P	2.4923	µg/l	6/7/2018 14:38:13	
J3350-01	S	Total CN	P	-0.8994	µg/l	6/7/2018 14:38:14	
CCV2	S	Total CN	P	246.0296	µg/l	6/7/2018 14:38:16	
CCB2	S	Total CN	P	-0.1452	µg/l	6/7/2018 14:38:17	
J3354-02	S	Total CN	P	-0.266	µg/l	6/7/2018 14:38:18	
J3354-02DUP	S	Total CN	P	-0.326	µg/l	6/7/2018 14:38:19	
J3354-03	S	Total CN	P	30.8902	µg/l	6/7/2018 14:38:20	
J3354-04	S	Total CN	P	35.0971	µg/l	6/7/2018 14:38:21	
J3354-05	S	Total CN	P	-0.1484	µg/l	6/7/2018 14:38:22	
J3352-01	S	Total CN	P	-0.2125	µg/l	6/7/2018 14:59:43	
CCV3	S	Total CN	P	259.9267	µg/l	6/7/2018 14:59:44	
CCB3	S	Total CN	P	-0.5483	µg/l	6/7/2018 14:59:45	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

6/7/2018 10:08


Test Total CN

Accepted 6/7/2018 10:08

Factor 968.7
 Bias 0.004

Coeff. of det. 0.999936

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.2083	0.0000	
2	5.0PPBCN	0.009	4.2290	5.0000	
3	10PPBCN	0.014	9.6063	10.0000	
4	50PPBCN	0.056	50.2503	50.0000	
5	100PPBCN	0.109	100.9506	100.0000	
6	250PPBCN	0.265	252.7433	250.0000	
7	500PPBCN	0.519	498.4288	500.0000	