


ISM02.4

LB 95880
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44630



A= 0.0000e+000

B= 2.2534e-004

C= -6.1368e-002

Rho= 0.9999590

Accept=Accepted

Slope
+1 m/intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.002	0.002	280	0.000	280				
0.2	0.200	0.171	-0.029	1033	0.0 %	1033				
2.5	2.500	2.560	0.060	11635	0.0 %	11635				
5.0	5.000	4.956	-0.044	22264	0.0 %	22264				
7.5	7.500	7.515	0.015	33622	0.0 %	33622				
10.0	10.000	9.996	-0.004	44630	0.0 %	44630				

0/0 D

-15

2

-1

0

0

NB

6/7/18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	280	-		0 ISM02.4	PPB	6/7/2018 17:25	Std
	0.2 S0.2	1033	-		0.2 ISM02.4	PPB	6/7/2018 17:27	Std
	2.5 S2.5	11635	-		2.5 ISM02.4	PPB	6/7/2018 17:30	Std
	5 S5	22264	-		5 ISM02.4	PPB	6/7/2018 17:32	Std
	7.5 S7.5	33622	-		7.5 ISM02.4	PPB	6/7/2018 17:36	Std
	10 S10.0	44630	-		10 ISM02.4	PPB	6/7/2018 17:43	Std
ICV	ICV	17297	3.8364 -		ISM02.4	PPB	6/7/2018 17:46	SMPL
ICB	ICB	106	-0.0375 -		ISM02.4	PPB	6/7/2018 17:48	SMPL
CCV079	CCV079	22204	4.9421 -		ISM02.4	PPB	6/7/2018 17:50	SMPL
CCB079	CCB079	99	-0.0391 -		ISM02.4	PPB	6/7/2018 17:53	SMPL
PB110053BL	PBW003	151	-0.0273 -		ISM02.4	PPB	6/7/2018 17:55	SMPL
J3367-01	MC0AB2	180	-0.0208 -		ISM02.4	PPB	6/7/2018 17:57	SMPL
J3367-02	MC0AB3	192	-0.0181 -		ISM02.4	PPB	6/7/2018 18:00	SMPL
J3367-03	MC0AB3D	166	-0.024 -		ISM02.4	PPB	6/7/2018 18:02	SMPL
J3367-04	MC0AB3S	4626	0.9811 -		ISM02.4	PPB	6/7/2018 18:04	SMPL
J3367-05	MC0AB4	121	-0.0341 -		ISM02.4	PPB	6/7/2018 18:06	SMPL
J3367-06	MC0AB5	141	-0.0296 -		ISM02.4	PPB	6/7/2018 18:09	SMPL
J3367-07	MC0AB6	129	-0.0323 -		ISM02.4	PPB	6/7/2018 18:11	SMPL
J3367-08	MC0AB7	115	-0.0355 -		ISM02.4	PPB	6/7/2018 18:13	SMPL
CCV080	CCV080	21754	4.8407 -		ISM02.4	PPB	6/7/2018 18:15	SMPL
CCB080	CCB080	101	-0.0386 -		ISM02.4	PPB	6/7/2018 18:18	SMPL