



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95887

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/07/2018	16:35
2	CAL2	1	Water			21.5	7.00	06/07/2018	16:37
3	CAL3	1	Water			21.7	10.03	06/07/2018	16:39
4	ICV	1	Water			21.4	7.00	06/07/2018	16:41
5	CCV1	1	Water			21.6	2.00	06/07/2018	16:43
6	J3240-02	1	Solid	20.02	20	22.5	6.27	06/07/2018	16:45
7	J3240-02DUP	1	Solid	20.04	20	22.6	6.28	06/07/2018	16:47
8	J3247-02	1	Solid	20.03	20	21.9	7.80	06/07/2018	16:49
9	J3360-03	1	Solid	20.02	20	21.3	7.05	06/07/2018	16:51
10	J3377-02	1	Solid	20.04	20	22.7	7.12	06/07/2018	16:53
11	J3377-05	1	Solid	20.05	20	22.6	6.50	06/07/2018	16:55
12	J3377-09	1	Solid	20.02	20	21.9	9.19	06/07/2018	16:57
13	CCV2	1	Water			21.5	12.00	06/07/2018	16:59

WORKLIST(Hardcopy Internal Chain)

LB95887

WorkList Name : corrosivity060718

WorkList ID : 113086

Date : 6/7/2018 4:19:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2018	Solid	J3240-02	Corrosivity	Cool 4 deg C	PSEG05	I63	BU-02-060718	06/07/2018	9045C
06/07/2018	Solid	J3247-02	Corrosivity	Cool 4 deg C	PSEG05	I62	OR-02-060718	06/07/2018	9045C
06/07/2018	Solid	J3360-03	Corrosivity	Cool 4 deg C	PSEG01	P42	JC-SOIL-COMP	06/07/2018	9045C
06/07/2018	Solid	J3377-02	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-B-DB-S	06/07/2018	9045C
06/07/2018	Solid	J3377-05	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-B-D	06/07/2018	9045C
06/07/2018	Solid	J3377-09	Corrosivity	Cool 4 deg C	PSEG01	I63	TP-C-DB	06/07/2018	9045C

Date/Time 06/07/18 4:25 PM

Received by: SP

Relinquished by: CP

Date/Time 06/07/18 5:10 PM

Received by: CP

Relinquished by: SP