


7470A

LB95889
INSTRUMENT ID: CV1

Linear

 μ Abs.:

3670E



A= 0.0000e+000

B= 2.6801e-004

C= -4.1686e-002

Rho= 0.9991797

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	137	0.000	137				
0.2	0.200	0.197	-0.003	890	0.0 %	890				
2.5	2.500	2.461	-0.039	9337	0.0 %	9337				
5.0	5.000	4.953	-0.047	18635	0.0 %	18635				
7.5	7.500	7.798	0.298	29251	0.0 %	29251				
10.0	10.000	9.797	-0.203	36708	0.0 %	36708				

LB95889 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jun 2018 10:05:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	137	-	0.00007470A	PPB		08 Jun 2018 10:08:23	S	Std
0.2 - 1	50.2	890	-	0.20007470A	PPB		08 Jun 2018 10:10:32	S	Std
2.5 - 1	52.5	9337	-	2.50007470A	PPB		08 Jun 2018 10:12:40	S	Std
5.0 - 1	55	18635	-	5.00007470A	PPB		08 Jun 2018 10:14:48	S	Std
7.5 - 1	57.5	29251	-	7.50007470A	PPB		08 Jun 2018 10:16:56	S	Std
10.0 - 1	51.5	36708	-	10.00007470A	PPB		08 Jun 2018 10:19:04	S	Std
ICV90 - 1	ICV90	14183	3.7596	-7470A	PPB		08 Jun 2018 10:21:48	U	SMPL
ICB90 - 1	ICB90	12	-0.0385	-7470A	PPB		08 Jun 2018 10:23:54	U	SMPL
CCV81 - 1	CCV81	18526	4.9235	-7470A	PPB		08 Jun 2018 10:26:01	U	SMPL
CCB81 - 1	CCB81	-6	-0.0433	-7470A	PPB		08 Jun 2018 10:28:08	U	SMPL
CRA - 1	CRA	873	0.1923	-7470A	PPB		08 Jun 2018 10:30:15	U	SMPL
HighStd - 1	HighStd	36915	9.8521	-7470A	PPB		08 Jun 2018 10:32:22	U	SMPL
ChkStd - 1	ChkStd	26805	7.1424	-7470A	PPB		08 Jun 2018 10:34:29	U	SMPL
PB110058BL - 1	PBW	-55	-0.0564	-7470A	PPB		08 Jun 2018 10:36:38	U	SMPL
PB110058BS - 1	LCSW	13373	3.5425	-7470A	PPB		08 Jun 2018 10:38:47	U	SMPL
PB110058TB - 1	PB110023TB	24	-0.0353	-7470A	PPB		08 Jun 2018 10:40:55	U	SMPL
J3252-02 - 1	TR-04-060618	130	-0.0068	-7470A	PPB		08 Jun 2018 10:43:03	U	SMPL
J3252-02DUP - 1	TR-04-060618DUP	146	-0.0026	-7470A	PPB		08 Jun 2018 10:45:11	U	SMPL
J3252-02MS - 1	TR-04-060618MS	14444	3.8295	-7470A	PPB		08 Jun 2018 10:47:20	U	SMPL
J3252-02MSD - 1	TR-04-060618MSD	13457	3.5650	-7470A	PPB		08 Jun 2018 10:49:27	U	SMPL
CCV82 - 1	CCV82	18575	4.9367	-7470A	PPB		08 Jun 2018 10:51:33	U	SMPL
CCB82 - 1	CCB82	-5	-0.0430	-7470A	PPB		08 Jun 2018 10:53:40	U	SMPL
J3344-04 - 1	HS-YARD-1	110	-0.0122	-7470A	PPB		08 Jun 2018 10:55:47	U	SMPL
J3344-08 - 1	HS-YARD-2	122	-0.0090	-7470A	PPB		08 Jun 2018 10:57:54	U	SMPL
J3344-12 - 1	HS-HOUSE-WALL	128	-0.0074	-7470A	PPB		08 Jun 2018 11:00:00	U	SMPL
J3352-04 - 1	PS-SOIL	130	-0.0068	-7470A	PPB		08 Jun 2018 11:02:07	U	SMPL
PB110059BL - 1	PBW	123	-0.0087	-7470A	PPB		08 Jun 2018 11:04:15	U	SMPL
PB110059BS - 1	LCSW	14140	3.7480	-7470A	PPB		08 Jun 2018 11:06:22	U	SMPL
J3330-01 - 1	CAS1-GW-01	524	0.0988	-7470A	PPB		08 Jun 2018 11:08:30	U	SMPL
J3330-01DUP - 1	CAS1-GW-01DUP	402	0.0661	-7470A	PPB		08 Jun 2018 11:10:39	U	SMPL
J3330-01MS - 1	CAS1-GW-01MS	15960	4.2358	-7470A	PPB		08 Jun 2018 11:12:48	U	SMPL
J3330-01MSD - 1	CAS1-GW-01MSD	15658	4.1549	-7470A	PPB		08 Jun 2018 11:14:54	U	SMPL
CCV83 - 1	CCV83	18571	4.9356	-7470A	PPB		08 Jun 2018 11:17:02	U	SMPL
CCB83 - 1	CCB83	4	-0.0406	-7470A	PPB		08 Jun 2018 11:19:09	U	SMPL
J3252-02LX5 - 1		124	-0.0085	-7470A	PPB		08 Jun 2018 11:21:17	U	SMPL
J3252-02A - 1		14959	3.9675	-7470A	PPB		08 Jun 2018 11:23:23	U	SMPL
J3330-01LX5 - 1		93	-0.0168	-7470A	PPB		08 Jun 2018 11:25:30	U	SMPL
J3330-01A - 1		16559	4.3964	-7470A	PPB		08 Jun 2018 11:27:37	U	SMPL
CCV84 - 1	CCV84	18869	5.0155	-7470A	PPB		08 Jun 2018 11:29:45	U	SMPL
CCB84 - 1	CCB84	-3	-0.0425	-7470A	PPB		08 Jun 2018 11:31:54	U	SMPL