

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB95899

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP65153
calibration std. phosphate 0.5 ppm	WP65152
calibration std. phosphate 0.3 ppm	WP65151
calibration std. 0 ppm	WP65148
phosphate ICV std.	WP65155
phosphate CCV std.	WP65154
calibration std. phosphate 0.1 ppm	WP65150
calibration std. phosphate 0.05 ppm	WP65149
Combined reagent	WP65160
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/08/2018	11:32
2	CAL2	0.05	1	50	50	0.033	06/08/2018	11:32
3	CAL3	0.10	1	50	50	0.065	06/08/2018	11:33
4	CAL4	0.30	1	50	50	0.186	06/08/2018	11:33
5	CAL5	0.50	1	50	50	0.330	06/08/2018	11:34
6	CAL6	1.00	1	50	50	0.649	06/08/2018	11:34



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB95899

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.495	06/08/2018	11:35
2	ICB		1	50	50	0.000	0.001	06/08/2018	11:35
3	CCV1	0.50	1	50	50	0.325	0.501	06/08/2018	11:36
4	CCB1		1	50	50	0.000	0.001	06/08/2018	11:36
5	PB110087BL		1	50	50	0.000	0.001	06/08/2018	11:37
6	PB110087BS	0.50	1	50	50	0.310	0.478	06/08/2018	11:37
7	J3208-01		1	50	50	0.028	0.044	06/08/2018	11:38
8	J3208-02		1	50	50	0.033	0.052	06/08/2018	12:38
9	J3208-02DUP		1	50	50	0.036	0.057	06/08/2018	12:39
10	J3208-02MS	0.50	1	50	50	0.348	0.536	06/08/2018	12:39
11	J3208-02MSD	0.50	1	50	50	0.350	0.540	06/08/2018	11:40
12	CCV2	0.50	1	50	50	0.327	0.504	06/08/2018	11:40
13	CCB2		1	50	50	0.000	0.001	06/08/2018	11:41