

LB95900


7471B

INSTRUMENT ID: CY1

Linear

μ Abs.:

44455



A= 0.0000e+000

B= 2.2494e-004

C= -2.8313e-002

Rho= 0.9999632

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	79	0.000	79				
0.2	0.200	0.200	0.000	1015	0.0 %	1015				
2.5	2.500	2.471	-0.029	11109	0.0 %	11109				
5.0	5.000	5.065	0.065	22644	0.0 %	22644				
7.5	7.500	7.502	0.002	33479	0.0 %	33479				
10.0	10.000	9.972	-0.028	44459	0.0 %	44459				

LB95900 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Jun 2018 12:04:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.0000	7471B	PPB	08 Jun 2018 12:08:15	S	Std
0.2 - 1	50.2	1015	-	0.2000	7471B	PPB	08 Jun 2018 12:10:32	S	Std
2.5 - 1	52.5	11109	-	2.5000	7471B	PPB	08 Jun 2018 12:12:50	S	Std
5.0 - 1	55	22644	-	5.0000	7471B	PPB	08 Jun 2018 12:15:07	S	Std
7.5 - 1	57	33479	-	7.5000	7471B	PPB	08 Jun 2018 12:17:25	S	Std
10.0 - 1	57	44459	-	10.0000	7471B	PPB	08 Jun 2018 12:19:42	S	Std
ICV91 - 1	ICV91	17411	3.8881	-7471B	PPB	PPB	08 Jun 2018 12:22:23	U	SMPL
ICB91 - 1	ICB91	-53	-0.0402	-7471B	PPB	PPB	08 Jun 2018 12:24:39	U	SMPL
CCV85 - 1	CCV85	22771	5.0938	-7471B	PPB	PPB	08 Jun 2018 12:26:56	U	SMPL
CCB85 - 1	CCB85	-73	-0.0447	-7471B	PPB	PPB	08 Jun 2018 12:29:11	U	SMPL
CRA - 1	CRA	1000	0.1966	-7471B	PPB	PPB	08 Jun 2018 12:31:28	U	SMPL
HighStd - 1	HighStd	45216	10.1426	-7471B	PPB	PPB	08 Jun 2018 12:33:44	U	SMPL
ChkStd - 1	ChkStd	30832	6.9070	-7471B	PPB	PPB	08 Jun 2018 12:35:59	U	SMPL
PB110066BL - 1	PBS	-136	-0.0589	-7471B	PPB	PPB	08 Jun 2018 12:38:18	U	SMPL
PB110066BS - 1	LCSS	15691	3.5012	-7471B	PPB	PPB	08 Jun 2018 12:40:37	U	SMPL
J3214-03 - 1	MS-CON-2	911	0.1766	-7471B	PPB	PPB	08 Jun 2018 12:42:54	U	SMPL
J3240-01 - 1	BU-02-060718	1035	0.2045	-7471B	PPB	PPB	08 Jun 2018 12:45:11	U	SMPL
J3247-01 - 1	OR-02-060718	1787	0.3737	-7471B	PPB	PPB	08 Jun 2018 12:47:29	U	SMPL
J3354-02 - 1	CAS1-SB-02-03-1-27	4105	0.8951	-7471B	PPB	PPB	08 Jun 2018 12:49:46	U	SMPL
J3354-02DUP - 1	CAS1-SB-02-03-1-27DUP	4002	0.8719	-7471B	PPB	PPB	08 Jun 2018 12:54:49	U	SMPL
CCV86 - 1	CCV86	22566	5.0477	-7471B	PPB	PPB	08 Jun 2018 12:57:04	U	SMPL
CCB86 - 1	CCB86	-74	-0.0450	-7471B	PPB	PPB	08 Jun 2018 12:59:21	U	SMPL
J3354-03 - 1	J3354-02MS	21373	4.7793	-7471B	PPB	PPB	08 Jun 2018 13:01:37	U	SMPL
J3354-04 - 1	J3354-02MSD	20952	4.6846	-7471B	PPB	PPB	08 Jun 2018 13:03:53	U	SMPL
J3354-05 - 1	CAS1-SB-102-103-1-27	736	0.1372	-7471B	PPB	PPB	08 Jun 2018 13:06:10	U	SMPL
J3360-01 - 1	JC-SOIL-COMP	2712	0.5817	-7471B	PPB	PPB	08 Jun 2018 13:08:27	U	SMPL
J3374-01 - 1	EPE-109-1(74-76)	653	0.1186	-7471B	PPB	PPB	08 Jun 2018 13:10:44	U	SMPL
J3374-02 - 1	EPE-109-1(80-82)	1043	0.2063	-7471B	PPB	PPB	08 Jun 2018 13:13:00	U	SMPL
J3374-03 - 1	EPE-109-1(85-87)	774	0.1458	-7471B	PPB	PPB	08 Jun 2018 13:15:17	U	SMPL
J3377-01 - 1	TP-B-DB-S	773	0.1456	-7471B	PPB	PPB	08 Jun 2018 13:17:35	U	SMPL
J3377-04 - 1	TP-B-D	224	0.0221	-7471B	PPB	PPB	08 Jun 2018 13:19:52	U	SMPL
J3377-08 - 1	TP-C-DB	2889	0.6215	-7471B	PPB	PPB	08 Jun 2018 13:22:08	U	SMPL
CCV87 - 1	CCV87	22965	5.1374	-7471B	PPB	PPB	08 Jun 2018 13:24:25	U	SMPL
CCB87 - 1	CCB87	-69	-0.0438	-7471B	PPB	PPB	08 Jun 2018 13:26:41	U	SMPL
J3354-02LX5 - 1		794	0.1503	-7471B	PPB	PPB	08 Jun 2018 13:28:58	U	SMPL
J3354-02A - 1		21618	4.8344	-7471B	PPB	PPB	08 Jun 2018 13:31:14	U	SMPL
CCV88 - 1	CCV88	22875	5.1172	-7471B	PPB	PPB	08 Jun 2018 13:33:30	U	SMPL
CCB88 - 1	CCB88	-79	-0.0461	-7471B	PPB	PPB	08 Jun 2018 13:35:47	U	SMPL